

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & KS Q ISO/IEC 17025:2017

Q & Q Co., Ltd.

58, Hwajeonsadan 3-ro, Gangseo-gu, Busan, Republic of Korea  
Phone : 82-51-292-0395~6, Fax : 82-51-292-0397, e-mail : qnq0395@hanmail.net

CALIBRATION

Valid To : Dec. 02, 2026.

Accreditation No : KC01-079

In recognition of the successful completion of the KOLAS evaluation process,  
accreditation is granted to this laboratory to perform the following calibrations

Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site
102. Linear dimension			10323	Alignment telescopes	N	10609	Micro indicators, test indicators	Y
10201	Balls	N	10326	Laser level	N	10610	Micrometer heads	Y
10206	Dial/cylinder gauge testers	N	10327	Optical wedges	N	10611	3-points, micrometers	Y
10207	Doctor blades	N	104. Form			10612	Inside micrometers	Y
10208	Laser distance meters	N	10401	Form testers	Y	10613	Outside micrometers	Y
10209	End bars	N	10404	Optical flats	N	10617	Standard sieves	N
10210	Extensometers	Y	10405	Optical parallels	N	10620	Welding gauges	Y
10211	Feeler gauges	Y	10406	Parallel blocks	N	10621	Optical micrometers	N
10212	Film applicators	N	10407	Precision surface plates	Y	201. Mass		
10213	Gap gauges	N	10409	Roundness measurement instruments	Y	20109	Electric balances	Y
10214	Gauge blocks, by comparison	N	10412	Straight edges	N	20112	Platform scale balances	Y
10216	Height gauges/measuring machines	Y	10413	Straight rules	N	20113	Spring scale balances	Y
10220	Standard measuring machines	Y	10415	Test bars	N	20116	Weights	N
10223	Electronic micrometers	N	105. Complex geometry			202. Force		
10224	Height micrometers, riser blocks	N	10501	Base gauges for electric bulb	N	20203	Tension/compression testing machines	Y
			10502	Bench centers	N	20204	Push-pull gauges	N
10225	Laser scan micrometers	Y	10503	Contact coordinate measuring machines	Y	203. Torque		
10227	Standard tape rules	N	10504	Non-contact coordinate measuring machines	Y	20303	Torque wrenches/drivers	Y
10228	Cylindrical plug/pin gauges, thread measuring wire gauges	N	10511	Measuring microscopes, profile projectors	Y	204. Pressure		
10229	Radius gauges	N	10512	Micro measuring microscopes	N	20402	Manometers	N
10230	Cylindrical ring gauges	N	10514	Taper plug gauges	N	20404	Hydraulic pressure ballances	N
10232	Step gauges	N	10515	Taper ring gauges	N	20406	Absolute pressure gauges	Y
10233	Taper thickness gauges	N	10517	Stylus type roughness testers	Y	20407	Blood pressure gauges	Y
10234	Ultrasonic thickness gauges	Y	10518	Socket gauges for electric blub	N	20408	Compound pressure gauges	Y
10235	Ultrasonic/coating thickness specimens	N	10519	Roughness standard/comparison specimens	N	20409	Differential pressure gauges	Y
10236	Coating thickness testers	Y				20411	Gauge pressure gauges	Y
103. Angle						20412	Pressure transducers/transmitters	Y
10303	Autocollimators	N	10525	Thread plug gauges	N	20413	Dial type vacuum gauges	Y
10304	Bevel protractors	N	10526	Taper thread plug gauges	N	20414	Water depth meters	Y
10306	Clinometers	N	10527	Thread ring gauges	N	209. Fluid flow		
10308	Fine angle generators, level comparators	N	10528	Taper thread ring gauges	N	20908	Gas flowmeters; differential pressure	N
10310	Indexing tables	N	10529	V-blocks, box blocks	N	20909	Liquid flowmeters; differential pressure	N
10311	Plate/square/electric levels	N	106. Various dimensional			20910	Liquid flowmeters; electromagnetic	N
10312	Auto levels	N	10601	Inside/outside/gear tooth calipers, caliper gauges	Y	20911	Gas flowmeters; thermal mass, etc.	N
10316	Rotary tables	N	10603	Cylinder/bore gauges	Y	20912	Liquid flowmeters; Coriolis, etc.	N
10317	Sine bars, plates, tables, centers	N	10604	Depth gauges, depth micrometers	Y			
10318	Squareness testers, right angle testers	N	10605	Dial/digital gauges	Y			
10319	Cylindrical squares	N	10608	Grind gauges	N	20914	Gas flowmeters; positive displacement	N
10320	Precision squares	N						

209. Fluid flow			40302	Clamp ammeters/voltmeters	Y	901. Chemical analysis			
20915	Liquid flowmeters; positive displacement	N	40303	AC voltage/current calibrators	Y	90101	Breath alcohol analyzers	N	
			40310	Power factor meters	Y	90103	Gas analyzers	Y	
20916	Gas flowmeters; turbine	N	40311	AC power meters	Y				
			40312	AC power supplies	Y				
20917	Liquid flowmeters; turbine	N	40313	Puncture/safety testers	Y				
			40318	AC voltmeters	Y				
20918	Gas flowmeters; ultrasonic	N	404. Other DC & LF Measurements						
			40410	Line frequency meters	Y				
20919	Liquid flowmeters; ultrasonic	N	40411	Function generators	Y				
			40414	LF impulse generators	Y				
20920	Gas flowmeters; variable area	N	40416	Leakage current testers	Y				
			40417	Electronic AC/DC loads	Y				
20921	Liquid flowmeters; variable area	N	40419	Analogue/Digital multimeters	Y				
			40421	Oscilloscopes	Y				
20922	Gas flowmeters; vortex	N	40424	Volt/current recorders	Y				
			40425	Relay test sets	Y				
20923	Liquid flowmeters; vortex	N	501. Contact thermometry						
			50101	Temperature generators; ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	Y				
210. Hardness			50102	Temperature indicators/ recorders/controllers, temperature calibrators					
21001	Brinell hardness testers	Y		50103 Glass thermometers; liquid -in-glass, Beckmann					
21002	Rockwell hardness testers	Y	50104	Resistance thermometers; SPRT, IPRT, thermistors, etc.					
21003	Shore hardness testers	Y		50105 Thermal expansion thermometers; bimetal, gas or liquid type					
21004	Vickers hardness testers	Y	50106	50106 Thermomecoules; noble metal, base metal					
21005	Durometer hardness testers	N		50107 Temperature transducers					
21006	Leeb hardness testers	N	50108	502. Non contact thermometry					
301. Time/frequency				50204 Standard radiation thermometers					
30103	General frequency sources	N	50109	401. DC voltage & current					
30104	Frequency meters/counters	N		503. Humidity					
30106	Time interval meters/ stop watches/timers	Y	50302	50302 Relative humidity hygrometers; polimer thin film, hair, etc.					
302. Velocity & revolution				50304 Temperature humidity recorders; hygrothermograph, etc.					
30201	Standard RPM generators	Y	50305	50305 Transducers; relative humidity					
30202	Contact type tachometers	N		50306 Humidity generators; two-pressure, two-temperature, flow mixing humidity gererator, constant temperature and humidity chamber, etc.					
30203	Photo tachometers /stroboscopes	N	701. Photometry						
402. Resistance, Capacitance and Inductance			70101	Iluminance meters	N				
40205	Earth testers	Y	40210	403. AC voltage, current & power					
40210	Insulation testers	Y		40301 AC ammeters					
40213	Resistance bridges & Similar instruments	N	40302	Note					
40214	Resistance meters	Y		1. This laboratory provides calibration services in permanent standard laboratory and at on-site.					
40215	Resistors	N	40303	2. Laboratory conducts on-site calibration should meet requirements of KOLAS-SR-007.					
403. AC voltage, current & power				3. On-site calibration is allowed to items with marking 'Y', not allowed to items with marking 'N'.					
4. Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k=2$ . It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.			40304	4. Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k=2$ . It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.					
5.Due to the calibration environment such as reference standards or customers' facilities, it is note that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.				5.Due to the calibration environment such as reference standards or customers' facilities, it is note that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.					

## 102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Balls Q & Q Co., Ltd.	10201	( $\phi$ 0 ~ $\phi$ 40) mm	$\sqrt{0.3^2 + 0.002 9^2 \times l^2} \mu\text{m}$	Standard measuring machine /QECI-LE201
		( $\phi$ 40 ~ $\phi$ 100) mm	$\sqrt{0.4^2 + 0.002 9^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	
Roundness		( $\phi$ 10 ~ $\phi$ 100) mm	0.2 $\mu\text{m}$	
Dial/cylinder gauge testers	10206	(0 ~ 5) mm (5 ~ 25) mm (25 ~ 100) mm	$\sqrt{0.11^2 + 0.002 7^2 \times l^2} \mu\text{m}$ $\sqrt{0.11^2 + 0.002 9^2 \times l^2} \mu\text{m}$ $\sqrt{0.21^2 + 0.002 9^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Gauge block /QECI-LE206
Doctor blades	10207	(0 ~ 10) mm	1.0 $\mu\text{m}$	Height micrometer /QECI-LE207
Laser distance meters	10208	(0 ~ 25) m	$\sqrt{0.8^2 + 0.001 5^2 \times l^2} \text{mm}$ (Unit of $l$ : m)	Laser Calibration System /QECI-LE208
End bars	10209	(25 ~ 1 000) mm (1 000 ~ 2 000) mm	$\sqrt{0.3^2 + 0.002 9^2 \times l^2} \mu\text{m}$ $\sqrt{1.2^2 + 0.003 0^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Gauge block, Electronic micrometer /QECI-LE209
Extensometers	10210	(0 ~ 500) mm	$\sqrt{1.2^2 + 0.002 8^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Gauge block /QECI-LE210
Feeler gauges	10211	(0 ~ 5) mm	0.3 $\mu\text{m}$	Standard measuring machine /QECI-LE211
Film applicators	10212	(0 ~ 10) mm	1 $\mu\text{m}$	Height micrometer /QECI-LE212
Gap gauges	10213	(1 ~ 200) mm (200 ~ 500) mm	$\sqrt{1.3^2 + 0.002 6^2 \times l^2} \mu\text{m}$ $\sqrt{2.0^2 + 0.004 7^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Height micrometer, Electronic micrometer /QECI-LE213
Gauge blocks, by comparison	10214	(0.5 ~ 100) mm (125 ~ 500) mm	$\sqrt{81^2 + 1.21^2 \times l^2} \text{nm}$ $\sqrt{95^2 + 1.30^2 \times l^2} \text{nm}$ (Unit of $l$ : mm)	Gauge block /QECI-LE214
Height gauges/measuring machines	10216			Gauge block /QECI-LE216
Height gauges		(0 ~ 1 500) mm	$\sqrt{7^2 + 0.002 9^2 \times l^2} \mu\text{m}$	
Height measuring machines		(0 ~ 1 000) mm	$\sqrt{0.6^2 + 0.002 9^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	

## 102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard measuring machines	10220	(0 ~ 500) mm	$\sqrt{0.2^2 + 0.0027^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Gauge block /QECI-LE220
Electronic micrometers	10223	$\pm 50 \mu\text{m}$ $\pm 50 \mu\text{m} \sim \pm 2 \text{ mm}$	0.1 $\mu\text{m}$ 1 $\mu\text{m}$	Gauge block /QECI-LE223
Height micrometers Block calibration	10224	(0 ~ 310) mm (310 ~ 1 010) mm	$\sqrt{0.6^2 + 0.0028^2 \times l^2} \mu\text{m}$ $\sqrt{0.7^2 + 0.0029^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Gauge block, Electronic micrometer /QECI-LE224, /QECI-LE224-1
Head calibration		(0 ~ 20) mm	0.6 $\mu\text{m}$	
Riser blocks		150 mm 300 mm 600 mm	0.8 $\mu\text{m}$ 1.0 $\mu\text{m}$ 1.8 $\mu\text{m}$	
Parallelism			0.6 $\mu\text{m}$	
Laser scan micrometers	10225	(ø0 ~ ø60) mm	$\sqrt{0.31^2 + 0.0038^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Cylindrical plug/pin gauge /QECI-LE225
Standard tape rules	10227	(0 ~ 25) m (25 ~ 50) m (50 ~ 75) m (75 ~ 100) m	$\sqrt{0.03^2 + 0.0015^2 \times l^2} \text{ mm}$ $\sqrt{0.05^2 + 0.0015^2 \times l^2} \text{ mm}$ $\sqrt{0.10^2 + 0.0015^2 \times l^2} \text{ mm}$ $\sqrt{0.12^2 + 0.0015^2 \times l^2} \text{ mm}$ (Unit of $l$ : m)	Laser tape measurement system /QECI-LE227
Cylindrical plug/pin gauges, thread measuring wire gauges	10228			Standard measuring machine,
Cylindrical plug/pin gauges		(ø0.2 ~ ø200) mm (ø200 ~ ø500) mm	$\sqrt{0.27^2 + 0.0037^2 \times l^2} \mu\text{m}$ $\sqrt{1.9^2 + 0.0048^2 \times l^2} \mu\text{m}$	Roundness measurement instrument
Thread measuring wire gauges		(ø0.17 ~ ø4.39) mm	$\sqrt{0.39^2 + 0.0036^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	/QECI-LE228
Roundness		(ø1 ~ ø200) mm	0.3 $\mu\text{m}$	
Radius gauges	10229	(0 ~ 100) mm (100 ~ 500) mm	$\sqrt{0.6^2 + 0.0028^2 \times l^2} \mu\text{m}$ $\sqrt{2.0^2 + 0.0041^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Non-contact coordinate measuring machine /QECI-LE229
Cylindrical ring gauges	10230	(ø2 ~ ø200) mm (ø200 ~ ø500) mm	$\sqrt{0.1^2 + 0.0028^2 \times l^2} \mu\text{m}$ $\sqrt{1.9^2 + 0.0047^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Gauge block, Standard measuring
Roundness		(ø2 ~ ø500) mm (ø2 ~ ø200) mm (ø200 ~ ø500) mm	0.2 $\mu\text{m}$ $\sqrt{0.3^2 + 0.0038^2 \times l^2} \mu\text{m}$ $\sqrt{2.7^2 + 0.0058^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	machine /QECI-LE230

## 102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Step gauges	10232	(0 ~ 300) mm (300 ~ 1 000) mm (1 000 ~ 1 500) mm	$\sqrt{1.2^2 + 0.002 7^2 \times l^2} \mu\text{m}$ $\sqrt{1.2^2 + 0.002 8^2 \times l^2} \mu\text{m}$ $\sqrt{1.6^2 + 0.002 8^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Gauge block /QECI-LE232
Taper thickness gauges	10233	(0 ~ 90) mm	$\sqrt{1.4^2 + 0.003 7^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Non-contact coordinate measuring machine /QECI-LE233
Ultrasonic thickness gauges	10234	(0 ~ 500) mm	$\sqrt{8^2 + 0.003 9^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Ultrasonic specimen /QECI-LE234
Ultrasonic/coating thickness specimens ; coating Ultrasonic Flatness	10235	(0 ~ 30) mm (2 ~ 800) mm	$\sqrt{0.3^2 + 0.002 7^2 \times l^2} \mu\text{m}$ $\sqrt{0.7^2 + 0.002 9^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm) 0.5 $\mu\text{m}$	Standard measuring machine /QECI-LE235, /QECI-LE235-1
Coating thickness testers	10236	(0 ~ 20) mm	$\sqrt{1.2^2 + 0.002 7^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Thickness specimen /QECI-LE236

## 103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Autocollimators	10303	$\pm 15'$	$\sqrt{0.4^2 + 0.000 3^2 \times R^2} "$ (Unit of $R$ : " )	Level comparator /QECI-AN303
Bevel protractors Angle accuracy Straightness	10304	(0 ~ 360) $^\circ$	1' 1.2 $\mu\text{m}$	Rotary table, Electronic micrometer /QECI-AN304
Clinometers	10306	(0 ~ 360) $^\circ$	4 "	Rotary table /QECI-AN306
Fine angle generators, Level comparators Angle accuracy Flatness	10308	$\pm 15'$ 530 mm $\times$ 60 mm	$\sqrt{0.6^2 + 0.000 3^2 \times R^2} "$ (Unit of $R$ : " ) 1.3 $\mu\text{m}$	Autocollimator /QECI-AN308
Indexing tables	10310	(0 ~ 360) $^\circ$	1.0 "	Autocollimator, Polygon /QECI-AN310

## 103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Plate/square/electric levels Angle(Bubble Tube Type)	10311	$\pm 0.1 \text{ mm/m}$ $(\pm 10 \sim \pm 20) \text{ mm/m}$	$\sqrt{0.52^2 + 0.000 3^2 \times R^2}''$ 0.6 mm/m	Level comparator Rotary Table /QECI-AN311,
Angle(Electric Type)		$\pm 5 \text{ mm/m}$ $(\pm 5 \sim \pm 10) \text{ mm/m}$	$\sqrt{0.3^2 + 0.000 3^2 \times R^2}''$	/QECI-AN311-1
Flatness of Base Squareness		$(0 \sim 300) \text{ mm}$ $(0 \sim 300) \text{ mm}$	$\sqrt{1.5^2 + 0.000 3^2 \times R^2}''$ (Unit of $R$ : '') 0.9 $\mu\text{m}$ 6.3 $\mu\text{m}/\text{m}$	
Auto levels Azimuth angle	10312	$(0 \sim 360)^\circ$	3 ''	Collimating calibration system
Line of sight straightness Optical Micrometer		0 m ~ $\infty$ $\pm 2.5 \text{ mm}$	0.15 mm 3 $\mu\text{m}$	/QECI-AN312
Rotary tables	10316	$(0 \sim 360)^\circ$	1.0 ''	Autocollimator, Polygon /QECI-AN316
Sine bars, plates, tables, Centers Distance between center of roller	10317	$(50 \sim 200) \text{ mm}$	1.8 $\mu\text{m}$	Angle gauge block, Electronic micrometer /QECI-AN317
Flatness of measuring surface			0.5 $\mu\text{m}$	
Parallelism between the measuring surface and the roller			0.5 $\mu\text{m}$	
Squareness testers	10318	$(0 \sim 480) \text{ mm}$	$\sqrt{1.4^2 + 0.003^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Standard cylindrical square /QECI-AN318
Cylindrical squares	10319	$(0 \sim 500) \text{ mm}$	1.5 $\mu\text{m}$	Standard cylindrical square /QECI-AN319
Precision squares Squareness	10320	$(0 \sim 500) \text{ mm}$	$\sqrt{1.3^2 + 0.003^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Cylindrical square, Contact coordinate measuring machine
Parallelism		$(500 \sim 1\,000) \text{ mm}$	4.0 $\mu\text{m}$	/QECI-AN320
Straightness		$(0 \sim 1\,000) \text{ mm}$	1.5 $\mu\text{m}$	
		$(0 \sim 1\,000) \text{ mm}$	1.9 $\mu\text{m}$	
Alignment telescopes Angle accuracy	10323	$\pm 5'$	1.3 ''	Collimator /QECI-AN323
Line of sight Straightness Optical Micrometer		0 m ~ $\infty$ $\pm 2.5 \text{ mm}$	0.15 mm 3 $\mu\text{m}$	

## 103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Laser levels	10326	(0 ~ 2) m	0.11 mm	CCD CAMERA /QECI-AN326
Horizontality Squareness	10326	90 °	0.07 °	
Optical wedges	10327	± 15 "	0.9 "	Autocollimator /QECI-AN327

## 104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Form testers	10401	(0 ~ 10) μm	0.04 μm	Step block, Gauge block, Pitch master /QECI-LE401
Height (Z axis)		(0.01 ~ 20) mm	0.07 μm	
Traversing length (X axis)		(0 ~ 50) mm	1.00 μm	Optical flat /QECI-LE404
Angle		1" ~ 45°	3 "	
Optical flats	10404	(ø 10 ~ ø 100) mm	$\sqrt{23^2 + 0.428^2 \times d^2}$ nm (Unit of d : mm)	Optical flat /QECI-LE404
Optical parallels	10405	(ø 10 ~ ø 50) mm	0.04 μm	Optical flat, Gauge block comparator /QECI-LE405
Flatness Parallelism			0.08 μm	
Parallel blocks	10406	(0 ~ 500) mm	0.8 μm	Electronic micrometer /QECI-LE406
Parallelism			0.8 μm	
Flatness			0.8 μm	
Difference of height between parallel block 1 and 2				
Precision surface plates	10407	(0.09 ~ 1) m <sup>2</sup> (1 ~ 1.44) m <sup>2</sup> (1.44 ~ 2.7) m <sup>2</sup> (2.7 ~ 4.84) m <sup>2</sup> (4.84 ~ 9) m <sup>2</sup> (9 ~ 17.5) m <sup>2</sup>	0.7 μm 0.9 μm 1.1 μm 1.4 μm 1.8 μm 2.2 μm	Electronic Level /QECI-LE407

## 104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Roundness measurement instruments	10409	(0 ~ 0.1) mm	0.30 $\mu\text{m}$ $2.6 \times 10^{-3}$	Roundness Standard Specimen, Optical flat, Cylindrical square /QECI-LE409
Detector accuracy		360°	0.03 $\mu\text{m}$	
Magnification accuracy		360°	0.03 $\mu\text{m}$	
Circumferential direction rotating error of spindle		(0 ~ 450) mm	1.3 $\mu\text{m}$	
Axial direction rotating error of spindle				
Straightness of column				
Straight edges	10412	(0 ~ 2 000) mm	0.6 $\mu\text{m}$ 1.5 $\mu\text{m}$	Electronic micrometer, Precision surface plate /QECI-LE412
Straight rules	10413	(0 ~ 5) m	$\sqrt{0.3^2 + 0.0015^2 \times l^2} \text{ mm}$ ( <i>l</i> 의 단위 : m)	Laser tape measurement system /QECI-LE413
Test bars	10415	(0 ~ 500) mm		Gauge block,
Outside diameter		(ø 10 ~ ø 100) mm	$\sqrt{0.4^2 + 0.0036^2 \times l^2} \mu\text{m}$ (Unit of <i>l</i> : mm)	Electronic micrometer /QECI-LE415
Straightness			0.9 $\mu\text{m}$	
Run-out			0.9 $\mu\text{m}$	
Flank angle			2 "	

## 105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Base gauges for electric bulb Inside diameter of GO/NO-GO side Screw	10501	( ø 1 ~ ø 50) mm ( ø 1 ~ ø 50) mm	$\sqrt{0.2^2 + 0.003 7^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm) $\sqrt{0.9^2 + 0.004 0^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Height micrometer, Form tester /QECI-LE501
Bench centers Height difference of both the center Flatness of the bed side	10502	(0 ~ 500) mm	2.4 $\mu\text{m}$ 1.1 $\mu\text{m}$	Test bar, Electronic micrometer /QECI-LE502
Contact coordinate measuring machines Indicating accuracy Squareness Straightness	10503	(0 ~ 1 000) mm (1 000 ~ 1 500) mm	$\sqrt{1.9^2 + 0.004 0^2 \times l^2} \mu\text{m}$ $\sqrt{2.0^2 + 0.004 0^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm) 4.0 $\mu\text{m}$ 0.9 $\mu\text{m}$	Gauge block, Precision square, Straight edge /QECI-LE503
Non-contact coordinate measuring machines Indicating accuracy Angle Squareness	10504	(0 ~ 600) mm (0 ~ 180) ° (0 ~ 450) mm	$\sqrt{0.4^2 + 0.002 8^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm) 5 ° 0.2 °	Standard scale, Angle gauge block /QECI-LE504
Measuring microscopes, Profile projectors Feed accuracy of workstage Squareness Magnification Error Angle division accuracy	10511	(0 ~ 300) mm	$\sqrt{0.4^2 + 0.002 8^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm) 1.7 $\mu\text{m}$ $2.4 \times 10^{-4}$ 1.1 ′	Standard scale, Square /QECI-LE511, /QECI-LE511-1
Micro measuring microscopes	10512	(0 ~ 20) mm	0.7 $\mu\text{m}$	Standard scale /QECI-LE512
Taper plug gauges Taper half angle Small-end diameter Step diameter Big-end diameter Gauge length Step length	10514	(0 ~ 65) ° ( ø 2 ~ ø 200) mm ( ø 2 ~ ø 200) mm ( ø 2 ~ ø 200) mm (0 ~ 250) mm (0 ~ 150) mm	1 ° $\sqrt{0.6^2 + 0.002 7^2 \times l^2} \mu\text{m}$ $\sqrt{0.6^2 + 0.002 7^2 \times l^2} \mu\text{m}$ $\sqrt{0.7^2 + 0.003 0^2 \times l^2} \mu\text{m}$ $\sqrt{0.6^2 + 0.003 0^2 \times l^2} \mu\text{m}$ $\sqrt{0.6^2 + 0.003 0^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Standard measuring machine, Electronic micrometer /QECI-LE514

## 105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Taper ring gauges	10515	(0 ~ 65)°	1 "	Standard measuring machine, Electronic micrometer, Contact coordinate measuring machine /QECI-LE515	
Taper half angle		(ø 2 ~ ø 80) mm	$\sqrt{0.5^2 + 0.0006^2 \times l^2} \mu\text{m}$		
Small-end diameter		(ø 80 ~ ø 200) mm	$\sqrt{1.9^2 + 0.0041^2 \times l^2} \mu\text{m}$		
Step diameter		(ø 2 ~ ø 80) mm	$\sqrt{0.5^2 + 0.0006^2 \times l^2} \mu\text{m}$		
		(ø 80 ~ ø 200) mm	$\sqrt{2.0^2 + 0.0041^2 \times l^2} \mu\text{m}$		
Big-end diameter		(ø 2 ~ ø 80) mm	$\sqrt{0.5^2 + 0.0006^2 \times l^2} \mu\text{m}$		
		(ø 80 ~ ø 200) mm	$\sqrt{2.3^2 + 0.0041^2 \times l^2} \mu\text{m}$		
Gauge length		(0 ~ 250) mm	$\sqrt{0.6^2 + 0.0030^2 \times l^2} \mu\text{m}$		
Notch & Step length		(0 ~ 150) mm	$\sqrt{0.6^2 + 0.0030^2 \times l^2} \mu\text{m}$		
			(Unit of $l$ : mm)		
Stylus type roughness testers	10517	(0 ~ 3) µm	$\sqrt{0.096^2 + 0.025^2 \times R^2} \mu\text{m}$	Roughness standard specimen, Step block /QECI-LE517	
Arithmetic mean(Ra)		(3 ~ 10) µm	$\sqrt{0.18^2 + 0.025^2 \times R^2} \mu\text{m}$		
Max. height(Rz)		(0 ~ 10) µm	$\sqrt{0.26^2 + 0.030^2 \times R^2} \mu\text{m}$		
		(10 ~ 50) µm	$\sqrt{0.27^2 + 0.003^2 \times R^2} \mu\text{m}$		
		(0 ~ 140) µm	$\sqrt{0.87^2 + 0.0036^2 \times L^2} \mu\text{m}$		
		(140 ~ 230) µm	$\sqrt{1.5^2 + 0.0036^2 \times L^2} \mu\text{m}$		
Mean width(RSm)		(0 ~ 1) µm	$\sqrt{0.02^2 + 0.019^2 \times H^2} \mu\text{m}$		
		(1 ~ 1 000) µm	$\sqrt{0.08^2 + 0.019^2 \times H^2} \mu\text{m}$		
			(Unit of $H$ : µm)		
			(Unit of $L$ : µm)		
Depth(H)		(0 ~ 1) µm	$\sqrt{0.022^2 + 0.019^2 \times H^2} \mu\text{m}$		
		(1 ~ 1 000) µm	$\sqrt{0.089^2 + 0.019^2 \times H^2} \mu\text{m}$		
Socket gauges for electric bulb	10518	(ø 1 ~ ø 50) mm	$\sqrt{0.3^2 + 0.0037^2 \times l^2} \mu\text{m}$	Form tester /QECI-LE518	
Outside diameter of GO/NOT GO /Thread GO side			(Unit of $l$ : mm)		
Roughness standard/ comparison specimens	10519	(0 ~ 3) µm	$\sqrt{0.12^2 + 0.025^2 \times R^2} \mu\text{m}$	Stylus type roughness tester /QECI-SR519	
Arithmetic mean(Ra)		(3 ~ 10) µm	$\sqrt{0.19^2 + 0.025^2 \times R^2} \mu\text{m}$		
Max. height(Rz)		(0 ~ 10) µm	$\sqrt{0.37^2 + 0.030^2 \times R^2} \mu\text{m}$		
		(10 ~ 50) µm	$\sqrt{0.29^2 + 0.003^2 \times R^2} \mu\text{m}$		
		(0 ~ 140) µm	$\sqrt{0.99^2 + 0.0036^2 \times L^2} \mu\text{m}$		
		(140 ~ 230) µm	$\sqrt{1.5^2 + 0.0036^2 \times L^2} \mu\text{m}$		
		(0 ~ 1) µm	$\sqrt{0.022^2 + 0.019^2 \times H^2} \mu\text{m}$		
		(1 ~ 1 000) µm	$\sqrt{0.089^2 + 0.019^2 \times H^2} \mu\text{m}$		
			(Unit of $H$ : µm)		
			(Unit of $L$ : µm)		
			(Unit of $R$ : µm)		

## 105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Thread plug gauges				
Pitch diameter	10525	( $\phi 1 \sim \phi 200$ ) mm	$\sqrt{1.3^2 + 0.003 7^2 \times l^2} \mu\text{m}$	Standard measuring machine,
Major diameter		( $\phi 1 \sim \phi 200$ ) mm	$\sqrt{0.4^2 + 0.003 7^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Form tester /QECI-LE525
Pitch		(0.25 ~ 6.35) mm	0.4 $\mu\text{m}$	
Flank angle		(0 ~ 45) °	4 "	
Taper thread plug gauges	10526	( $\phi 2 \sim \phi 200$ ) mm	$\sqrt{1.6^2 + 0.004 3^2 \times M_0^2} \mu\text{m}$ (Unit of $M_0$ : mm)	Standard measuring machine,
Small-end pitch diameter		( $\phi 2 \sim \phi 200$ ) mm	$\sqrt{1.6^2 + 0.003 4^2 \times M_H^2} \mu\text{m}$ (Unit of $M_H$ : mm)	Gauge block /QECI-LE526
Large-end pitch diameter		( $\phi 2 \sim \phi 200$ ) mm	$\sqrt{0.7^2 + 0.004 3^2 \times L_0^2} \mu\text{m}$ (Unit of $L_0$ : mm)	
Small-end major diameter		( $\phi 2 \sim \phi 200$ ) mm	$\sqrt{0.7^2 + 0.003 4^2 \times L_H^2} \mu\text{m}$ (Unit of $L_H$ : mm)	
Large-end major diameter		( $\phi 2 \sim \phi 200$ ) mm	0.9 $\mu\text{m}$	
Pitch		(0.25 ~ 6.35) mm	5 "	
Flank angles		(0 ~ 30) °	7 "	
Taper half-angle		(0 ~ 2) °		
Gauge length		(0 ~ 250) mm	$\sqrt{0.6^2 + 0.003 0^2 \times l^2} \mu\text{m}$	
Notch & Step length		(0 ~ 150) mm	$\sqrt{0.6^2 + 0.003 0^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	
Thread ring gauges	10527	( $\phi 2.5 \sim \phi 200$ ) mm	$\sqrt{1.2^2 + 0.003 7^2 \times l^2} \mu\text{m}$	Standard measuring machine,
Pitch diameter		( $\phi 2.5 \sim \phi 200$ ) mm	$\sqrt{1.7^2 + 0.005 6^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Form tester /QECI-LE527
Minor diameter		(0.25 ~ 6.35) mm	1.0 $\mu\text{m}$	
Pitch		(0 ~ 45) °	5 "	
Flank angle				
Taper thread ring gauges	10528	(1 ~ 150) mm	$\sqrt{0.6^2 + 0.003 0^2 \times l^2} \mu\text{m}$	Height micrometer,
Gauge length		(0 ~ 150) mm	$\sqrt{0.6^2 + 0.003 0^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Electronic micrometer,
Notch & Step length		(0 ~ 2) mm	1.2 $\mu\text{m}$	Taper thread plug gauge /QECI-LE528
Alternateness of minor diameter		(0 ~ 2) mm	2.1 $\mu\text{m}$	
Alternateness of pitch diameter				

## 105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
V-blocks, Box blocks	10529	300 mm × 300 mm × 300 mm		Electronic micrometer, Test bar,
V-blocks			0.5 μm	Squareness tester
Flatness of base side			0.6 μm	/QECI-LE529,
Flatness of V surface			0.4 μm	/QECI-LE529-1
The gradient on the base side of V-groove				
The parallelism between the undersurface and the cylinder on the V surface.			2.4 μm	
The parallelism between the side and the cylinder on the V surface.			2.4 μm	
The mutual height difference of V surface for a pair of V blocks			2.3 μm	
Box blocks			1.6 μm	
Squareness			1.0 μm	
The parallelism of upper surface for the undersurface				
The parallelism between the undersurface and the cylinder on the V surface.			2.6 μm	

## 106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inside/outside/gear tooth calipers, caliper gauges Inside/Outside calipers	10601	(0 ~ 600) mm (600 ~ 1 000) mm (1 000 ~ 3 000) mm	$\sqrt{8^2 + 0.002 9^2 \times l^2} \mu\text{m}$ $\sqrt{10^2 + 0.002 9^2 \times l^2} \mu\text{m}$ $\sqrt{15^2 + 0.002 9^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Gauge block, Surface plate /QECI-LE601, /QECI-LE601-1, /QECI-LE601-2
Gear tooth calipers Tooth height scale Tooth thickness scale Combine accuracy Caliper gauges		(0 ~ 50) mm (0 ~ 100) mm (0 ~ 100) mm (0 ~ 300) mm	$\sqrt{7^2 + 0.002 7^2 \times l^2} \mu\text{m}$ $\sqrt{7^2 + 0.002 9^2 \times l^2} \mu\text{m}$ $\sqrt{7^2 + 0.002 6^2 \times l^2} \mu\text{m}$ $\sqrt{7^2 + 0.002 9^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	
Cylinder/Bore gauges	10603	(0 ~ 5) mm	0.3 $\mu\text{m}$	Standard measuring machine, Dial gauge tester /QECI-LE603
Depth gauges, Depth micrometers Depth gauges Depth micrometers Dial depth gauges	10604	(0 ~ 1 000) mm (0 ~ 300) mm (0 ~ 100) mm	$\sqrt{7^2 + 0.002 7^2 \times l^2} \mu\text{m}$ $\sqrt{1.0^2 + 0.002 7^2 \times l^2} \mu\text{m}$ $\sqrt{1.0^2 + 0.002 7^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Gauge block, Surface plate /QECI-LE604, /QECI-LE604-1, /QECI-LE604-2
Dial/digital gauges	10605	(0 ~ 5) mm (5 ~ 100) mm	$\sqrt{0.2^2 + 0.002 7^2 \times l^2} \mu\text{m}$ $\sqrt{0.8^2 + 0.002 7^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Standard measuring machine, Dial gauge tester /QECI-LE605
Grind gauges Depth of inclined plane Straightness of scraper	10608	(0 ~ 1) mm	0.9 $\mu\text{m}$ 1.1 $\mu\text{m}$	Height micrometer, Electronic micrometer /QECI-LE608
Micro indicators, Test indicators Micro indicators Test indicators	10609	$\pm 1$ mm (0 ~ 2) mm	0.3 $\mu\text{m}$ 0.3 $\mu\text{m}$	Dial gauge tester /QECI-LE609
Micrometer heads	10610	(0 ~ 100) mm	$\sqrt{0.2^2 + 0.002 9^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Gauge blocks /QECI-LE610
3-points, Micrometers	10611	( $\phi$ 2 ~ $\phi$ 6) mm ( $\phi$ 6 ~ $\phi$ 200) mm ( $\phi$ 200 ~ $\phi$ 300) mm	$\sqrt{1.4^2 + 0.004 6^2 \times l^2} \mu\text{m}$ $\sqrt{1.9^2 + 0.004 6^2 \times l^2} \mu\text{m}$ $\sqrt{2.5^2 + 0.005 4^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Cylindrical ring gauge /QECI-LE611

## 106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inside micrometers	10612	(5 ~ 300) mm	$\sqrt{1.0^2 + 0.0029^2 \times l^2} \mu\text{m}$	Gauge block,
Inside micrometers		(50 ~ 1 000) mm	$\sqrt{1.0^2 + 0.0029^2 \times l^2} \mu\text{m}$	Gauge block accessory
Tubular inside micrometers		(1 000 ~ 5 000) mm	$\sqrt{2.0^2 + 0.0029^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	/QECI-LE612
Outside micrometers	10613	(0 ~ 1 000) mm	$\sqrt{1.0^2 + 0.0029^2 \times l^2} \mu\text{m}$	Gauge block,
Outside micrometers		(1 000 ~ 2 000) mm	$\sqrt{2.0^2 + 0.0029^2 \times l^2} \mu\text{m}$	Optical flat ,
V-Anvil micrometers		(1 ~ 100) mm	$\sqrt{1.0^2 + 0.0049^2 \times l^2} \mu\text{m}$ (Unit of $l$ : mm)	Cylindrical plug gauge /QECI-LE613, /QECI-LE613-1
Standard sieves	10617			Non-contact coordinate measuring machine
Standard net sieve		(0 ~ 10) mm	3 $\mu\text{m}$	/QECI-LE617
Wire diameter		(0 ~ 130) mm	4 $\mu\text{m}$	
Sieve size				
Standard plate sieve		(0 ~ 130) mm	3 $\mu\text{m}$	
Hole diameter		(0 ~ 160) mm	3 $\mu\text{m}$	
Length of the hole center				
Welding gauges	10620			Non-contact coordinate measuring machine
Height, Depth		(0 ~ 50) mm	0.3 mm	/QECI-LE620
Scale		(0 ~ 90) mm	0.3 mm	
Fillet Welding Height		(0 ~ 20) mm	0.3 mm	
Taper thickness		(1 ~ 10) mm	0.2 mm	
Angle		(0 ~ 90) °	0.3 °	
Optical micrometers	10621	(0 ~ 10) mm	2 $\mu\text{m}$	Standard measuring machine /QECI-LE621

## 201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electric balances	20109	(0 ~ 20) g	64 µg	Weights /QECI-EB109
		(20 ~ 80) g	0.11 mg	
		(80 ~ 160) g	0.16 mg	
		(160 ~ 200) g	0.19 mg	
		(200 ~ 300) g	0.23 mg	
		(300 ~ 400) g	0.25 mg	
		(400 ~ 500) g	0.50 mg	
		(500 ~ 600) g	0.56 mg	
		(0.6 ~ 1) kg	0.9 mg	
		(1 ~ 2) kg	1.8 mg	
		(2 ~ 3) kg	1.9 mg	
		(3 ~ 4) kg	2.5 mg	
		(4 ~ 5) kg	4.7 mg	
		(5 ~ 8) kg	5.2 mg	
		(8 ~ 10) kg	9 mg	
		(10 ~ 16) kg	11 mg	
		(16 ~ 25) kg	18 mg	
		(25 ~ 30) kg	19 mg	
		(30 ~ 40) kg	0.02 g	
		(40 ~ 60) kg	0.7 g	
		(60 ~ 150) kg	2 g	
		(150 ~ 300) kg	3 g	
		(300 ~ 1 000) kg	0.1 kg	
		(1 000 ~ 2 000) kg	0.2 kg	
		(2 000 ~ 5 000) kg	0.5 kg	
Platform scale balances	20112	(0 ~ 10) kg	0.002 kg	Weights /QECI-EB112
		(10 ~ 20) kg	0.005 kg	
		(20 ~ 100) kg	0.01 kg	
		(100 ~ 200) kg	0.02 kg	
		(200 ~ 300) kg	0.05 kg	
		(300 ~ 500) kg	0.1 kg	
		(500 ~ 1 000) kg	0.23 kg	
Spring scale balances	20113	(0 ~ 500) g	1 g	Weights /QECI-EB113
		(0.5 ~ 1) kg	2 g	
		(1 ~ 2) kg	5 g	
		(2 ~ 5) kg	0.01 kg	
		(5 ~ 10) kg	0.02 kg	
		(10 ~ 30) kg	0.05 kg	
		(30 ~ 50) kg	0.1 kg	
		(50 ~ 100) kg	0.2 kg	

## 201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Weights, Class F1	20116	1 mg	2.2 $\mu$ g	Weights /QECI-EB116
		2 mg	2.2 $\mu$ g	
		5 mg	2.2 $\mu$ g	
		10 mg	3.0 $\mu$ g	
		20 mg	3.8 $\mu$ g	
		50 mg	4.2 $\mu$ g	
		100 mg	5.4 $\mu$ g	
		200 mg	6.8 $\mu$ g	
		500 mg	8.4 $\mu$ g	
		1 g	11 $\mu$ g	
		2 g	13 $\mu$ g	
		5 g	17 $\mu$ g	
		10 g	20 $\mu$ g	
		20 g	27 $\mu$ g	
		50 g	0.04 mg	
		100 g	0.06 mg	
		200 g	0.14 mg	
		500 g	0.31 mg	
		1 kg	0.55 mg	
		2 kg	1.9 mg	
		5 kg	2.7 mg	
		10 kg	5.7 mg	
		20 kg	10 mg	

## 202. Force

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Tension/compression testing machines	20203	Tension	(2 ~ 20) kN	Load Cells /QECI-FC203
			(20 ~ 50) kN	
			(2 ~ 10) kN	
			(10 ~ 20) kN	
			(20 ~ 50) kN	
		Compression	(50 ~ 100) kN	
			(100 ~ 500) kN	
			(0.5 ~ 1) MN	
			(1 ~ 2) MN	

## 202. Force

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Push-pull gauges Tension, Compression	20204	(0.2 ~ 2) N	$1.5 \times 10^{-3}$	Weights /QECI-FC204
		(2 ~ 5) N	$8.0 \times 10^{-4}$	
		(5 ~ 10) N	$6.0 \times 10^{-4}$	
		(10 ~ 50) N	$5.0 \times 10^{-4}$	
		(50 ~ 100) N	$6.0 \times 10^{-4}$	
		(100 ~ 300) N	$5.0 \times 10^{-4}$	
		(300 ~ 1 000) N	$6.0 \times 10^{-4}$	

## 203. Torque

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Torque wrenches/drivers	20303	(0.1 ~ 1) N·m	$6.9 \times 10^{-3}$	Torque wrench tester /QECI-FC303
		(1 ~ 2.5) N·m	$9.8 \times 10^{-3}$	
		(2.5 ~ 5) N·m	$5.9 \times 10^{-3}$	
		(5 ~ 10) N·m	$7.3 \times 10^{-3}$	
		(10 ~ 25) N·m	$7.2 \times 10^{-3}$	
		(25 ~ 50) N·m	$3.6 \times 10^{-3}$	
		(50 ~ 100) N·m	$8.3 \times 10^{-3}$	
		(100 ~ 250) N·m	$1.8 \times 10^{-3}$	
		(250 ~ 500) N·m	$7.2 \times 10^{-3}$	
		(500 ~ 1 000) N·m	$6.2 \times 10^{-3}$	
		(1 000 ~ 2 000) N·m	$4.4 \times 10^{-3}$	

## 204. Pressure

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Manometers	20402	(0 ~ 1.25) kPa	$\sqrt{2.8^2 + 0.68^2} \times p_s^2 \text{ Pa}$ (Unit of $p_s$ : kPa)	Pressure Controller/Calibrator /QECI-PS402
		(1.25 ~ 15) kPa	$\sqrt{2.8^2 + 0.63^2} \times p_s^2 \text{ Pa}$ (Unit of $p_s$ : kPa)	
		(15 ~ 130) kPa	$\sqrt{77^2 + 0.11^2} \times p_s^2 \text{ Pa}$ (Unit of $p_s$ : kPa)	
Hydraulic pressure balances	20404	(0.1 ~ 10) MPa	$6.4 \times 10^{-5}$	Hydraulic pressure balance /QECI-PS404, /QECI-PS404-1
		(10 ~ 100) MPa	$6.6 \times 10^{-5}$	
		(100 ~ 200) MPa	$6.8 \times 10^{-5}$	
		(200 ~ 500) MPa	$1.7 \times 10^{-4}$	

## 204. Pressure

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Absolute pressure gauges Absolute pressure Barometers	20406	(0 ~ 350) kPa	$\sqrt{0.58^2 + 0.047^2 \times p_s^2}$ Pa ( $p_s$ : Max Pressure, Unit of $p_s$ : kPa)	Pressure Controller/Calibrator /QECI-PS406
		(0.35 ~ 5) MPa	$\sqrt{0.058^2 + 0.047^2 \times p_s^2}$ kPa ( $p_s$ : Max Pressure, Unit of $p_s$ : MPa)	
		(90 ~ 110) kPa	$\sqrt{9.0^2 + 0.15^2 \times p_s^2}$ Pa ( $p_s$ : Max Pressure, Unit of $p_s$ : kPa)	
Blood pressure gauges	20407	(0 ~ 40) kPa	$\sqrt{7.7^2 + 0.28^2 \times p_s^2}$ Pa ( $p_s$ : Max Pressure, Unit of $p_s$ : kPa)	Pressure Controller/Calibrator /QECI-PS407
Compound pressure gauges	20408	-100 kPa ~ 5 MPa	$\sqrt{0.058^2 + 0.066^2 \times p_s^2}$ kPa ( $p_s$ : the negative or positive max pressure, unit of $p_s$ : MPa)	Pressure Controller/Calibrator /QECI-PS408
Differential pressure gauges	20409	(0 ~ 1.25) kPa	$\sqrt{0.58^2 + 0.12^2 \times p_s^2}$ Pa ( $p_s$ : Max Pressure, Unit of $p_s$ : kPa)	Pressure Controller/Calibrator, Pneumatic pressure balance /QECI-PS409
		(1.25 ~ 15) kPa	$\sqrt{0.58^2 + 0.047^2 \times p_s^2}$ Pa ( $p_s$ : Max Pressure, Unit of $p_s$ : kPa)	
		(0.015 ~ 7) MPa	$\sqrt{0.058^2 + 0.049^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
Gauge pressure gauges Pneumatic Hydraulic	20411	(0 ~ 15) kPa	$\sqrt{0.10^2 + 0.047^2 \times p_s^2}$ Pa (Unit of $p_s$ : kPa)	Pressure Controller/Calibrator, Hydraulic pressure balance /QECI-PS411
		(0.015 ~ 0.7) MPa	$\sqrt{0.0058^2 + 0.053^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(0.7 ~ 7) MPa	$\sqrt{0.058^2 + 0.049^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(0 ~ 10) MPa	$\sqrt{0.080^2 + 0.060^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(10 ~ 100) MPa	$\sqrt{0.080^2 + 0.062^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(100 ~ 200) MPa	$\sqrt{0.080^2 + 0.063^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(200 ~ 500) MPa	$\sqrt{8.4^2 + 0.17^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	

## 204. Pressure

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Pressure transducers/transmitters	20412			Pressure Controller/Calibrator, Pneumatic pressure balance, Hydraulic pressure balance /QECI-PS412
Pressure transducers		(-100 ~ 0) kPa	$\sqrt{1.3^2 + 0.056^2 \times p_s^2}$ Pa ( $p_s$ : the negative maximum pressure, unit of $p_s$ : kPa)	
Pneumatic		(0 ~ 15) kPa	$\sqrt{0.20^2 + 0.047^2 \times p_s^2}$ Pa (Unit of $p_s$ : kPa)	
		(0.015 ~ 0.7) MPa	$\sqrt{0.0082^2 + 0.053^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(0.7 ~ 7) MPa	$\sqrt{0.090^2 + 0.049^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
Hydraulic		(0 ~ 10) MPa	$\sqrt{0.13^2 + 0.060^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(10 ~ 100) MPa	$\sqrt{1.3^2 + 0.062^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(100 ~ 200) MPa	$\sqrt{2.3^2 + 0.063^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(200 ~ 500) MPa	$\sqrt{6.5^2 + 0.17^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
Pressure transmitters		(-100 ~ 0) kPa	$\sqrt{3.8^2 + 0.056^2 \times p_s^2}$ Pa ( $p_s$ : the negative maximum pressure, unit of $p_s$ : kPa)	
Pneumatic		(0 ~ 15) kPa	$\sqrt{0.20^2 + 0.047^2 \times p_s^2}$ Pa (Unit of $p_s$ : kPa)	
		(0.015 ~ 0.7) MPa	$\sqrt{0.0084^2 + 0.053^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(0 ~ 7) MPa	$\sqrt{0.10^2 + 0.049^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
Hydraulic		(0 ~ 10) MPa	$\sqrt{0.15^2 + 0.060^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(10 ~ 100) MPa	$\sqrt{1.5^2 + 0.062^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(100 ~ 200) MPa	$\sqrt{3.0^2 + 0.063^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
		(200 ~ 500) MPa	$\sqrt{7.4^2 + 0.17^2 \times p_s^2}$ kPa (Unit of $p_s$ : MPa)	
Dial type vacuum gauges	20413	(-100 ~ 0) kPa	$\sqrt{0.10^2 + 0.000050^2 \times p_s^2}$ kPa ( $p_s$ : the negative maximum pressure, unit of $p_s$ : kPa)	Pressure Controller/Calibrator /QECI-PS413

## 204. Pressure

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Water Depth Meters	20414	(0 ~ 350) m (350 ~ 690) m	$\sqrt{0.088^2 + 0.25^2 \times p_s^2}$ m (Unit of $p_s$ : MPa) $\sqrt{0.41^2 + 0.25^2 \times p_s^2}$ m (Unit of $p_s$ : MPa)	Pressure Controller/Calibrator /QECI-PS414

## 209. Fluid flow

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Gas flowmeters; differential pressure	20908	(0.002 ~ 250) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Sonic nozzles /QECI-FL901
Liquid flowmeters; differential pressure	20909	(0.2 ~ 1 500) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Master Flow Meters /QECI-FL902
Liquid flowmeters; electromagnetic	20910	(0.2 ~ 1 500) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Master Flow Meters /QECI-FL902
Gas flowmeters; thermal mass, etc.	20911	(0.002 ~ 250) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Sonic nozzles /QECI-FL901
Liquid flowmeters; Coriolis, etc.	20912	(0.2 ~ 1 500) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Master Flow Meters /QECI-FL902
Gas flowmeters; positive displacement	20914	(0.002 ~ 250) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Sonic nozzles /QECI-FL901
Liquid flowmeters; positive displacement	20915	(0.2 ~ 1 500) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Master Flow Meters /QECI-FL902
Gas flowmeters; turbine	20916	(0.002 ~ 250) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Sonic nozzles /QECI-FL901
Liquid flowmeters; turbine	20917	(0.2 ~ 1 500) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Master Flow Meters /QECI-FL902
Gas flowmeters; ultrasonic	20918	(0.002 ~ 250) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Sonic nozzles /QECI-FL901
Liquid flowmeters; ultrasonic	20919	(0.2 ~ 1 500) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Master Flow Meters /QECI-FL902
Gas flowmeters; variable area	20920	(0.002 ~ 250) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Sonic nozzles /QECI-FL901
Liquid flowmeters; variable area	20921	(0.2 ~ 1 500) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Master Flow Meters /QECI-FL902
Gas flowmeters; vortex	20922	(0.002 ~ 250) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Sonic nozzles /QECI-FL901
Liquid flowmeters; vortex	20923	(0.2 ~ 1 500) m <sup>3</sup> /h	$3.2 \times 10^{-3}$	Master Flow Meters /QECI-FL902

## 210. Hardness

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Brinell hardness testers	21001	≤ 250 HBW 10/3 000 (250 ~ 450) HBW 10/3 000 > 450 HBW 10/3 000	2.6 HBW 10/3 000 4.1 HBW 10/3 000 6.0 HBW 10/3 000	Brinell Hardness test block /QECI-HN101
Rockwell hardness testers	21002	(20 ~ 70) HRC (10 ~ 100) HRBW	0.4 HRC 0.7 HRBW	Rockwell Hardness test block /QECI-HN102
Shore hardness testers	21003	(20 ~ 35) HS (45 ~ 55) HS (55 ~ 65) HS (75 ~ 85) HS (85 ~ 100) HS	1.0 HS 1.0 HS 1.0 HS 1.1 HS 1.1 HS	Shore hardness test block /QECI-HN103
Vickers hardness testers	21004	≤ 225 HV 0.2 (400 ~ 600) HV 0.2 > 700 HV 0.2 ≤ 225 HV 0.5 (400 ~ 600) HV 0.5 > 700 HV 0.5 ≤ 225 HV 10 (400 ~ 600) HV 10 > 700 HV 10 ≤ 225 HV 30 (400 ~ 600) HV 30 > 700 HV 30	5.7 HV 0.2 14 HV 0.2 20 HV 0.2 5.2 HV 0.5 15 HV 0.5 18 HV 0.5 3.0 HV 10 6.9 HV 10 8.9 HV 10 3.5 HV 30 5.7 HV 30 11 HV 30	Vickers hardness test block /QECI-HN104
Durometer hardness testers	21005	(0 ~ 100) HDA (0 ~ 100) HDD	0.4 HDA 0.4 HDD	Durometer calibrator /QECI-HN105
Leeb hardness testers	21006	≤ 500 HLD (500 ~ 700) HLD > 700 HLD	5 HLD 5 HLD 5 HLD	Leeb hardness test block /QECI-HN106

## 301. Time/frequency

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
General frequency sources Reference oscillator	30103	10 MHz	$7.0 \times 10^{-12}$	GPS receiver, Frequency counter /QECI-TF103
Frequency meters/counters Input Frequency Reference oscillator	30104	10 MHz 10 MHz	$7.0 \times 10^{-12}$ $7.0 \times 10^{-12}$	GPS receiver, Frequency counters /QECI-TF104
Time interval meters/ stop watches/timers Stop watches Timers	30106	(0.1 ~ 86 400) s (0.1 ~ 3.0) s (3.0 ~ 10 000) s	$1.5 \times 10^{-7}$ $6.4 \times 10^{-4}$ $7.0 \times 10^{-4}$	Stopwatch calibrator GPS receiver, Frequency counter /QECI-TF106

## 302. Velocity &amp; revolution

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard RPM generators RPM RPM(Centrifuge)	30201	(30 ~ 4 000) min <sup>-1</sup> (60 ~ 5 000) min <sup>-1</sup> (5 000 ~ 30 000) min <sup>-1</sup>	0.05 min <sup>-1</sup> 0.1 min <sup>-1</sup> 1 min <sup>-1</sup>	GPS receiver, Frequency counter stroboscope / QECI-VR201
Contact type tachometers RPM	30202	(6 ~ 60) min <sup>-1</sup> (60 ~ 4 000) min <sup>-1</sup>	0.01 min <sup>-1</sup> 0.1 min <sup>-1</sup>	GPS receiver, RPM calibration system / QECI-VR202
Photo tachometers/stroboscopes RPM (Tachometer) RPM (Stroboscopes)	30203	(30 ~ 1 000) min <sup>-1</sup> (1 000 ~ 10 000) min <sup>-1</sup> (10 000 ~ 99 000) min <sup>-1</sup> (30 ~ 1 000) min <sup>-1</sup> (1 000 ~ 10 000) min <sup>-1</sup> (10 000 ~ 99 000) min <sup>-1</sup>	0.001 min <sup>-1</sup> 0.01 min <sup>-1</sup> 0.1 min <sup>-1</sup> 0.001 min <sup>-1</sup> 0.01 min <sup>-1</sup> 0.1 min <sup>-1</sup>	GPS receiver Frequency counter, Optical generator/detector / QECI-VR203

## 401. DC volatage &amp; current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC ammeters	40101	(±)		
		0 $\mu$ A	9.3 nA	Meter calibrator,
		(0 ~ 10) $\mu$ A	$4.8 \times 10^{-3}$	Current calibrator /QECI-EL101
		(10 ~ 100) $\mu$ A	$5.5 \times 10^{-4}$	
		(0.1 ~ 1) mA	$1.5 \times 10^{-4}$	
		(1 ~ 10) mA	$1.0 \times 10^{-4}$	
		(10 ~ 100) mA	$1.3 \times 10^{-4}$	
		(0.1 ~ 1) A	$2.0 \times 10^{-4}$	
		(1 ~ 10) A	$9.0 \times 10^{-4}$	
		(10 ~ 100) A	$4.7 \times 10^{-4}$	
DC voltage/current calibrators	40103	(±)		
		0 mV	0.12 $\mu$ V	Multimeter /QECI-EL103
		(0 ~ 10) mV	$1.3 \times 10^{-4}$	
		(10 ~ 100) mV	$4.2 \times 10^{-5}$	
		(0.1 ~ 1) V	$8.7 \times 10^{-6}$	
		(1 ~ 10) V	$8.7 \times 10^{-6}$	
		(10 ~ 100) V	$9.3 \times 10^{-6}$	
		(100 ~ 1 000) V	$9.7 \times 10^{-6}$	
		0 $\mu$ A	0.74 nA	
		(0 ~ 100) $\mu$ A	$1.1 \times 10^{-4}$	
Electrical temperature calibrators	40104	(±)		
		0 mV	1 $\mu$ V	Multimeter Meter calibrator /QECI-EL104
		(0.000 ~ 5.239) mV	$3.8 \times 10^{-4}$	
		(0 ~ 600) °C		
		(5.239 ~ 18.503) mV	$1.1 \times 10^{-4}$	
		(600 ~ 1 750) °C		
		0 mV	1 $\mu$ V	
		(0.000 ~ 5.584) mV	$3.6 \times 10^{-4}$	
		(0 ~ 600) °C		
		(5.584 ~ 20.877) mV	$9.6 \times 10^{-5}$	
		(600 ~ 1 750) °C		

## 401. DC volatage &amp; current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
B Type	40104	(1.792 ~ 6.786) mV (600 ~ 1 200) °C (6.786 ~ 13.820) mV (1 200 ~ 1 820) °C	$1.1 \times 10^{-3}$ $1.4 \times 10^{-4}$ $3.4 \times 10^{-4}$	Multimeter Meter calibrator /QECI-EL104
K Type		(-5.891 ~ 12.209) mV (-200 ~ 300) °C 0 mV (12.209 ~ 54.886) mV (300 ~ 1 372) °C	$1 \mu\text{V}$ $3.6 \times 10^{-5}$	
J Type		(-7.891 ~ 10.779) mV (-200 ~ 200) °C 0 mV (10.779 ~ 57.953) mV (200 ~ 1 000) °C	$2.5 \times 10^{-4}$ $1 \mu\text{V}$ $3.5 \times 10^{-5}$	
T Type		(-5.603 ~ 4.279) mV (-200 ~ 100) °C 0 mV (4.279 ~ 20.872) mV (100 ~ 400) °C	$3.6 \times 10^{-4}$ $1 \mu\text{V}$ $9.6 \times 10^{-5}$	
N Type		(-3.990 ~ 9.341) mV (-200 ~ 300) °C 0 mV (9.341 ~ 47.513) mV (300 ~ 1 300) °C	$2.5 \times 10^{-4}$ $1 \mu\text{V}$ $4.2 \times 10^{-5}$	
E Type		(-8.825 ~ 37.005) mV (-200 ~ 500) °C 0 mV (37.005 ~ 76.362) mV (500 ~ 1 000) °C	$2.3 \times 10^{-4}$ $1 \mu\text{V}$ $2.6 \times 10^{-5}$	
pt100(385) Type		(18.52 ~ 332.79) Ω (-200 ~ 660) °C	$3.0 \times 10^{-5}$	
pt100(3916) Type		(17.08 ~ 337.03) Ω (-200 ~ 660) °C	$3.0 \times 10^{-5}$	
Measurement		(±)		
S Type		0 mV (0.000 ~ 5.239) mV (5.239 ~ 18.174) mV	$2 \mu\text{V}$ $3.8 \times 10^{-4}$ $1.1 \times 10^{-4}$	

## 401. DC volatage &amp; current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
R Type	40104	0 mV	2 $\mu$ V	Multimeter
		(0.000 ~ 5.584) mV	$3.6 \times 10^{-4}$	Meter calibrator
		(5.584 ~ 20.489) mV	$9.8 \times 10^{-5}$	/QECI-EL104
		(1.792 ~ 6.786) mV	$1.1 \times 10^{-3}$	
		(6.786 ~ 13.820) mV	$1.4 \times 10^{-4}$	
		(-5.891 ~ 12.209) mV	$3.4 \times 10^{-4}$	
		0 mV	2 $\mu$ V	
		(12.209 ~ 54.886) mV	$5.5 \times 10^{-5}$	
		(-7.891 ~ 10.779) mV	$2.5 \times 10^{-4}$	
		0 mV	2 $\mu$ V	
		(10.779 ~ 57.953) mV	$5.2 \times 10^{-5}$	
		(-5.603 ~ 4.279) mV	$3.6 \times 10^{-4}$	
T Type	40104	0 mV	2 $\mu$ V	
		(4.279 ~ 20.872) mV	$9.6 \times 10^{-5}$	
		(-3.990 ~ 9.341) mV	$5.0 \times 10^{-4}$	
		0 mV	2 $\mu$ V	
		(9.341 ~ 47.513) mV	$6.3 \times 10^{-5}$	
		(-8.825 ~ 37.005) mV	$2.3 \times 10^{-4}$	
		0 mV	2 $\mu$ V	
		(37.005 ~ 76.373) mV	$4.1 \times 10^{-5}$	
		(18.52 ~ 332.79) $\Omega$	$6.0 \times 10^{-5}$	
		(17.08 ~ 337.03) $\Omega$	$5.9 \times 10^{-5}$	
DC power supplies	40108	( $\pm$ )		
		0 mV	0.60 $\mu$ V	Multimeter,
		(0 ~ 100) mV	$3.2 \times 10^{-5}$	Shunt
		(0.1 ~ 1) V	$2.0 \times 10^{-5}$	/QECI-EL108
		(1 ~ 10) V	$2.0 \times 10^{-5}$	
		(10 ~ 100) V	$2.1 \times 10^{-5}$	
		(100 ~ 1 000) V	$2.1 \times 10^{-5}$	
		0 mA	5.9 nA	
		(0 ~ 1) mA	$3.7 \times 10^{-5}$	
		(1 ~ 10) mA	$3.6 \times 10^{-5}$	
		(10 ~ 100) mA	$7.7 \times 10^{-5}$	
		(0.1 ~ 1) A	$2.6 \times 10^{-4}$	
DC current	40108	(1 ~ 10) A	$6.0 \times 10^{-4}$	
		(10 ~ 100) A	$2.5 \times 10^{-4}$	

## 401. DC volatage &amp; current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC voltmeters DC voltmeter	40112	(±)		
		0 mV	0.61 $\mu$ V	Meter calibrator /QECI-EL112
		(0 ~ 10) mV	$6.5 \times 10^{-4}$	
		(10 ~ 100) mV	$4.0 \times 10^{-5}$	
		(0.1 ~ 1) V	$1.3 \times 10^{-5}$	
		(1 ~ 10) V	$1.0 \times 10^{-5}$	
		(10 ~ 100) V	$1.3 \times 10^{-5}$	
		(100 ~ 1 000) V	$1.3 \times 10^{-5}$	

## 402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Earth testers AC voltage Resistors meters	40205	60 Hz		
		(0.2 ~ 1) V	$6.5 \times 10^{-4}$	Meter calibrator, Decade resistance box /QECI-EL205
		(1 ~ 10) V	$3.9 \times 10^{-4}$	
		(10 ~ 100) V	$4.0 \times 10^{-4}$	
		(0.2 ~ 1) $\Omega$	$1.0 \times 10^{-3}$	
		(1 ~ 10) $\Omega$	$5.8 \times 10^{-4}$	
		(10 ~ 100) $\Omega$	$5.8 \times 10^{-4}$	
		(0.1 ~ 1) k $\Omega$	$5.8 \times 10^{-4}$	
		(1 ~ 10) k $\Omega$	$5.8 \times 10^{-4}$	
Insulation testers DC voltage AC voltage Resistance	40210	(0.01 ~ 0.1) kV	$5.8 \times 10^{-2}$	Multimeter, Meter calibrator,
		(0.1 ~ 1) kV	$7.3 \times 10^{-4}$	High Resistor /QECI-EL210
		(1 ~ 5) kV	$6.3 \times 10^{-3}$	
		(5 ~ 10) kV	$6.2 \times 10^{-3}$	
		60 Hz		
		(4 ~ 10) V	$3.8 \times 10^{-4}$	
		(10 ~ 100) V	$3.0 \times 10^{-4}$	
		(100 ~ 1 000) V	$4.0 \times 10^{-4}$	
		1 k $\Omega$	0.35 $\Omega$	
		(1 ~ 10) k $\Omega$	$3.5 \times 10^{-4}$	
		(10 ~ 100) k $\Omega$	$3.5 \times 10^{-4}$	
		(0.1 ~ 1) M $\Omega$	$3.5 \times 10^{-4}$	
		(1 ~ 10) M $\Omega$	$3.6 \times 10^{-4}$	
		(10 ~ 100) M $\Omega$	$1.2 \times 10^{-3}$	
		(0.1 ~ 1) G $\Omega$	$2.3 \times 10^{-3}$	
		(1 ~ 10) G $\Omega$	$6.0 \times 10^{-3}$	
		(10 ~ 100) G $\Omega$	$1.2 \times 10^{-2}$	
		(100 ~ 1 000) G $\Omega$	$1.2 \times 10^{-2}$	

## 402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance bridges & Similar instruments Measuring Arm	40213	(20 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ	$2.5 \times 10^{-4}$ $6.1 \times 10^{-5}$ $3.5 \times 10^{-5}$ $2.4 \times 10^{-5}$ $1.9 \times 10^{-5}$ $1.9 \times 10^{-5}$ $1.9 \times 10^{-5}$ $2.9 \times 10^{-5}$ $8.0 \times 10^{-5}$	Multimeter /QECI-EL213
Resistance meters Resistor	40214	1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (100 ~ 1 000) GΩ	0.06 μΩ $6.0 \times 10^{-5}$ $6.0 \times 10^{-5}$ $3.0 \times 10^{-5}$ $2.0 \times 10^{-5}$ $2.0 \times 10^{-5}$ $2.0 \times 10^{-5}$ $2.0 \times 10^{-5}$ $2.0 \times 10^{-5}$ $4.0 \times 10^{-5}$ $3.0 \times 10^{-5}$ $3.0 \times 10^{-3}$ $6.0 \times 10^{-3}$ $1.2 \times 10^{-2}$ $1.2 \times 10^{-2}$	Resistor, Decade resistance box, Resistance meter /QECI-EL214
Decade resistance box, Resistor	40215	1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ	0.66 μΩ $6.5 \times 10^{-4}$ $2.9 \times 10^{-4}$ $1.5 \times 10^{-4}$ $1.7 \times 10^{-5}$ $2.4 \times 10^{-5}$ $1.9 \times 10^{-5}$ $1.9 \times 10^{-5}$ $1.9 \times 10^{-5}$ $2.9 \times 10^{-5}$ $8.0 \times 10^{-5}$ $6.5 \times 10^{-4}$ $6.5 \times 10^{-3}$	Meter calibrator, Multimeter /QECI-EL215

## 403. AC voltage, current &amp; power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC ammeters	40301	50 Hz ~ 1 kHz		Meter calibrator,
AC current		(2 ~ 10) $\mu$ A	$3.1 \times 10^{-2}$	Current calibrator /QECI-EL301
		(10 ~ 100) $\mu$ A	$3.1 \times 10^{-3}$	
		(0.1 ~ 1) mA	$4.4 \times 10^{-4}$	
		(1 ~ 10) mA	$3.8 \times 10^{-4}$	
		(10 ~ 100) mA	$4.2 \times 10^{-4}$	
		(0.1 ~ 1) A	$1.0 \times 10^{-3}$	
		(1 ~ 10) A	$3.8 \times 10^{-3}$	
		60 Hz		
		(10 ~ 100) A	$2.9 \times 10^{-3}$	
Clamp ammeters/voltmeters	40302	(20 ~ 100) mV	$8.5 \times 10^{-5}$	Meter calibrator,
DC voltage		(0.1 ~ 1) V	$3.1 \times 10^{-5}$	Coil,
		(1 ~ 10) V	$2.0 \times 10^{-5}$	Decade resistance box
		(10 ~ 100) V	$2.8 \times 10^{-5}$	/QECI-EL302
		(100 ~ 1 000) V	$2.8 \times 10^{-5}$	
DC current		(0.3 ~ 1) A	$2.4 \times 10^{-3}$	
		(1 ~ 10) A	$2.4 \times 10^{-3}$	
		(10 ~ 100) A	$2.4 \times 10^{-3}$	
		(100 ~ 1 000) A	$2.7 \times 10^{-3}$	
AC voltage		60 Hz		
		(30 ~ 100) mV	$5.0 \times 10^{-4}$	
		(0.1 ~ 1) V	$3.8 \times 10^{-4}$	
		(1 ~ 10) V	$3.8 \times 10^{-4}$	
		(10 ~ 100) V	$3.0 \times 10^{-4}$	
		(100 ~ 1 000) V	$4.0 \times 10^{-4}$	
AC current		60 Hz		
		(0.3 ~ 1) A	$2.9 \times 10^{-3}$	
		(1 ~ 10) A	$2.9 \times 10^{-3}$	
		(10 ~ 100) A	$2.9 \times 10^{-3}$	
		(100 ~ 1 000) A	$2.9 \times 10^{-3}$	
Resistance		(2 ~ 10) $\Omega$	$8.0 \times 10^{-5}$	
		(10 ~ 100) $\Omega$	$6.6 \times 10^{-5}$	
		(0.1 ~ 1) k $\Omega$	$6.6 \times 10^{-5}$	
		(1 ~ 10) k $\Omega$	$6.6 \times 10^{-5}$	
		(10 ~ 100) k $\Omega$	$6.6 \times 10^{-5}$	
		(0.1 ~ 1) M $\Omega$	$1.5 \times 10^{-4}$	
		(1 ~ 10) M $\Omega$	$2.7 \times 10^{-4}$	

## 403. AC voltage, current &amp; power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Multimeter Shunt /QECI-EL303
AC voltage		40 Hz ~ 1 kHz (2 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	8.2 × 10 <sup>-3</sup> 8.4 × 10 <sup>-4</sup> 2.8 × 10 <sup>-4</sup> 2.8 × 10 <sup>-4</sup> 2.8 × 10 <sup>-4</sup> 3.2 × 10 <sup>-4</sup>	
AC current		40 Hz~ 1 kHz (2 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (50 ~ 400) Hz (10 ~ 100) A	1.7 × 10 <sup>-3</sup> 1.6 × 10 <sup>-3</sup> 2.5 × 10 <sup>-3</sup> 2.6 × 10 <sup>-3</sup>	
Power factor meters	40310	60 Hz 1 0.9 0.8 0.7 0.6 0.5 0.4 0.3 0.2 0.1	1.4 × 10 <sup>-3</sup> 1.6 × 10 <sup>-3</sup> 2.3 × 10 <sup>-3</sup> 3.1 × 10 <sup>-3</sup> 4.7 × 10 <sup>-3</sup> 7.6 × 10 <sup>-3</sup> 9.5 × 10 <sup>-3</sup> 1.9 × 10 <sup>-2</sup> 2.9 × 10 <sup>-2</sup> 1.2 × 10 <sup>-1</sup>	Power calibrator /QECI-EL310

## 403. AC voltage, current &amp; power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(50 ~ 60) Hz		Power calibrator, Coil /QECI-EL311
		(0.4 ~ 1) V	$3.8 \times 10^{-4}$	
		(1 ~ 10) V	$3.8 \times 10^{-4}$	
		(10 ~ 100) V	$3.3 \times 10^{-4}$	
		(100 ~ 1 000) V	$4.0 \times 10^{-4}$	
		(50 ~ 60) Hz		
		(2 ~ 10) mA	$3.0 \times 10^{-3}$	
		(10 ~ 100) mA	$1.1 \times 10^{-3}$	
		(0.1 ~ 1) A	$2.0 \times 10^{-3}$	
		(1 ~ 10) A	$2.2 \times 10^{-3}$	
		(10 ~ 20) A	$1.9 \times 10^{-3}$	
		(50 ~ 60) Hz		
		(2 ~ 10) W	$1.2 \times 10^{-3}$	
		(10 ~ 50) W	$1.2 \times 10^{-3}$	
		(50 ~ 100) W	$1.2 \times 10^{-3}$	
Electric power	40312	(100 ~ 500) W	$1.2 \times 10^{-3}$	Multimeter, Shunt /QECI-EL312
		(0.5 ~ 1) kW	$1.2 \times 10^{-3}$	
		(1 ~ 5) kW	$1.2 \times 10^{-3}$	
		(5 ~ 10) kW	$5.2 \times 10^{-4}$	
		(10 ~ 20) kW	$4.0 \times 10^{-4}$	
		40 Hz ~ 1 kHz		
		(20 ~ 100) mV	$8.4 \times 10^{-4}$	
		(0.1 ~ 1) V	$2.8 \times 10^{-4}$	
		(1 ~ 10) V	$2.8 \times 10^{-4}$	
		(10 ~ 100) V	$2.8 \times 10^{-4}$	
AC power supplies	40312	(100 ~ 1 000) V	$3.2 \times 10^{-4}$	Multimeter, Shunt /QECI-EL312
		40 Hz ~ 1 kHz		
		(2 ~ 10) mA	$1.7 \times 10^{-3}$	
		(10 ~ 100) mA	$1.7 \times 10^{-3}$	
		(0.1 ~ 1) A	$1.9 \times 10^{-3}$	
		(1 ~ 10) A	$3.3 \times 10^{-3}$	
		(50 ~ 400) Hz		
		(10 ~ 100) A	$1.3 \times 10^{-3}$	

## 403. AC voltage, current &amp; power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Puncture/safety testers	40313	(0.2 ~ 1) kV	$1.0 \times 10^{-2}$	High voltage voltmeter,	
		(1 ~ 5) kV	$7.0 \times 10^{-3}$	Current calibrator ,	
		(5 ~ 10) kV	$1.7 \times 10^{-2}$	Multimeter,	
		(10 ~ 50) kV	$1.3 \times 10^{-2}$	Oscilloscope	
		(50 ~ 100) kV	$1.3 \times 10^{-2}$	/QECI-EL313	
		60 Hz			
		(0.2 ~ 1) kV	$2.0 \times 10^{-2}$		
		(1 ~ 5) kV	$1.4 \times 10^{-2}$		
		(5 ~ 10) kV	$2.5 \times 10^{-2}$		
		(10 ~ 50) kV	$1.6 \times 10^{-2}$		
Breaking DC current		(50 ~ 100) kV	$1.3 \times 10^{-2}$		
		(0.2 ~ 1) mA	$5.0 \times 10^{-3}$		
		(1 ~ 10) mA	$5.0 \times 10^{-3}$		
Breaking AC current		(10 ~ 100) mA	$5.0 \times 10^{-3}$		
		60 Hz			
		(0.2 ~ 1) mA	$1.0 \times 10^{-2}$		
Operating time		(1 ~ 10) mA	$5.5 \times 10^{-3}$		
		(10 ~ 100) mA	$1.0 \times 10^{-2}$		
		(0 ~ 30) s	$1.0 \times 10^{-2}$		
AC voltmeters	40318	50 Hz ~ 1 kHz		Meter calibrator	
		(1 ~ 10) mV	$1.0 \times 10^{-2}$	/QECI-EL318	
		(10 ~ 100) mV	$6.5 \times 10^{-4}$		
		(0.1 ~ 1) V	$2.7 \times 10^{-4}$		
		(1 ~ 10) V	$1.8 \times 10^{-4}$		
		(10 ~ 100) V	$1.8 \times 10^{-4}$		
		(100 ~ 1 000) V	$1.7 \times 10^{-4}$		
		(1 ~ 10) kHz			
		(1 ~ 10) mV	$1.0 \times 10^{-2}$		
		(10 ~ 100) mV	$6.5 \times 10^{-4}$		
		(0.1 ~ 1) V	$2.7 \times 10^{-4}$		
		(1 ~ 10) V	$1.8 \times 10^{-4}$		
		(10 ~ 100) V	$1.8 \times 10^{-4}$		
		(10 ~ 100) kHz			
		(1 ~ 10) mV	$3.1 \times 10^{-2}$		
		(10 ~ 100) mV	$2.5 \times 10^{-3}$		
		(0.1 ~ 1) V	$1.2 \times 10^{-3}$		
		(1 ~ 10) V	$4.7 \times 10^{-4}$		
		(10 ~ 100) V	$9.0 \times 10^{-4}$		

## 404. Other DC &amp; LF Measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Line frequency meters	40410	(45 ~ 100) Hz (100 ~ 1 000) Hz (1 000 ~ 10 000) Hz	0.02 Hz 0.2 Hz 2 Hz	Meter calibrator QECI-EL410
Function generators	40411	0.1 Hz ~ 100 MHz 10 Hz ~ 1 kHz 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (1 ~ 10) kHz 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (10 ~ 100) kHz 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V 100 kHz ~ 1 MHz 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V 1 V 40 Hz 100 Hz 1 kHz 10 kHz 100 kHz 1 MHz 10 MHz 10 mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 20) V 1 ns ~ 10 ms	5.8 × 10 <sup>-7</sup> 0.21 μV 2.1 × 10 <sup>-4</sup> 2.1 × 10 <sup>-4</sup> 1.1 × 10 <sup>-4</sup> 1.1 × 10 <sup>-4</sup> 1.1 × 10 <sup>-4</sup> 0.43 μV 4.3 × 10 <sup>-4</sup> 4.3 × 10 <sup>-4</sup> 2.7 × 10 <sup>-4</sup> 2.7 × 10 <sup>-4</sup> 2.7 × 10 <sup>-4</sup> 0.94 μV 9.4 × 10 <sup>-4</sup> 9.4 × 10 <sup>-4</sup> 7.0 × 10 <sup>-4</sup> 7.0 × 10 <sup>-4</sup> 7.0 × 10 <sup>-4</sup> 25 μV 2.5 × 10 <sup>-2</sup> 2.4 × 10 <sup>-2</sup> 2.4 × 10 <sup>-2</sup> 2.4 × 10 <sup>-2</sup> 1.4 × 10 <sup>-4</sup> 1.3 × 10 <sup>-4</sup> 1.3 × 10 <sup>-4</sup> 2.7 × 10 <sup>-4</sup> 7.0 × 10 <sup>-4</sup> 2.1 × 10 <sup>-3</sup> 2.1 × 10 <sup>-3</sup> 0.58 μV 5.8 × 10 <sup>-5</sup> 5.9 × 10 <sup>-5</sup> 5.8 × 10 <sup>-5</sup> 3.0 × 10 <sup>-5</sup> 6.2 × 10 <sup>-3</sup>	GPS Receiver , Universal Counter , Multimeter , Oscilloscope QECI-EL411
Amplitude Flatness				
DC Offset				
Rise/Fall Time				

## 404. Other DC &amp; LF Measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF impulse generators				
Impulse voltage	40414	0.1 kV (0.1 ~ 1) kV (1 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV	4.3 V $4.3 \times 10^{-2}$ $4.3 \times 10^{-2}$ $4.3 \times 10^{-2}$ $4.3 \times 10^{-2}$	Oscilloscope, High voltage tester /QECI-EL414
Pulse Width		(20 ns ~ 100 ms)	$2.2 \times 10^{-3}$	
Pulse Rising Time		(20 ns ~ 100 ms)	$2.2 \times 10^{-3}$	
Leakage current testers				
AC voltage	40416	60 Hz (20 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 600) V	$7.5 \times 10^{-4}$ $4.4 \times 10^{-4}$ $2.9 \times 10^{-4}$ $3.0 \times 10^{-4}$ $4.2 \times 10^{-4}$	Meter calibrator, Current calibrator /QECI-EL416
AC current		60 Hz (2 ~ 10) $\mu$ A (10 ~ 100) $\mu$ A (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA	$9.5 \times 10^{-3}$ $1.2 \times 10^{-3}$ $5.6 \times 10^{-3}$ $3.0 \times 10^{-3}$ $2.9 \times 10^{-3}$	
DC current		(2 ~ 10) $\mu$ A (10 ~ 100) $\mu$ A (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA	$4.7 \times 10^{-3}$ $6.5 \times 10^{-4}$ $3.7 \times 10^{-4}$ $4.7 \times 10^{-4}$ $4.8 \times 10^{-4}$	
Electronic AC/DC loads				
DC voltage	40417	(20 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	$9.0 \times 10^{-5}$ $4.7 \times 10^{-5}$ $2.7 \times 10^{-5}$ $3.3 \times 10^{-5}$ $3.3 \times 10^{-5}$	Meter calibrator, Current calibrator /QECI-EL417
DC current		(20 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 100) A	$1.2 \times 10^{-4}$ $1.2 \times 10^{-4}$ $1.4 \times 10^{-4}$ $2.5 \times 10^{-4}$	

## 404. Other DC &amp; LF Measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Analogue/Digital multimeters	40419	(±)			
		0 mV	0.70 $\mu$ V	Resistors,	
		(0 ~ 100) mV	$4.6 \times 10^{-5}$	Decade resistance box	
		(0.1 ~ 1) V	$1.5 \times 10^{-5}$	/QECI-EL419	
		(1 ~ 10) V	$1.2 \times 10^{-5}$		
		(10 ~ 100) V	$1.4 \times 10^{-5}$		
		(100 ~ 1 000) V	$1.4 \times 10^{-5}$		
		0 $\mu$ A	9.3 nA		
		(0 ~ 100) $\mu$ A	$5.3 \times 10^{-4}$		
		(0.1 ~ 1) mA	$1.1 \times 10^{-4}$		
DC current		(1 ~ 10) mA	$9.3 \times 10^{-5}$		
		(10 ~ 100) mA	$1.0 \times 10^{-4}$		
		(0.1 ~ 1) A	$1.9 \times 10^{-4}$		
		(1 ~ 20) A	$9.5 \times 10^{-4}$		
		50 Hz ~ 10 kHz			
		(20 ~ 100) mV	$6.5 \times 10^{-4}$		
AC voltage		(0.1 ~ 1) V	$2.7 \times 10^{-4}$		
		(1 ~ 10) V	$1.8 \times 10^{-4}$		
		(10 ~ 100) V	$1.8 \times 10^{-4}$		
		(100 ~ 1 000) V	$1.7 \times 10^{-4}$		
		(10 ~ 100) kHz			
AC voltage		(20 ~ 100) mV	$2.5 \times 10^{-3}$		
		(0.1 ~ 1) V	$1.2 \times 10^{-3}$		
		(1 ~ 10) V	$4.7 \times 10^{-4}$		
		(10 ~ 100) V	$9.3 \times 10^{-4}$		
AC current		50 Hz ~ 1 kHz			
		(20 ~ 100) $\mu$ A	$2.9 \times 10^{-4}$		
		(0.1 ~ 1) mA	$4.4 \times 10^{-4}$		
		(1 ~ 10) mA	$2.7 \times 10^{-4}$		
		(10 ~ 100) mA	$2.8 \times 10^{-4}$		
		(0.1 ~ 1) A	$1.0 \times 10^{-3}$		
Resistance		(1 ~ 20) A	$2.3 \times 10^{-3}$		
		1 $\Omega$	24 $\mu$ $\Omega$		
		(1 ~ 10) $\Omega$	$1.2 \times 10^{-5}$		
		(10 ~ 100) $\Omega$	$1.2 \times 10^{-5}$		
		(0.1 ~ 1) k $\Omega$	$1.2 \times 10^{-5}$		
		(1 ~ 10) k $\Omega$	$1.2 \times 10^{-5}$		
		(10 ~ 100) k $\Omega$	$1.2 \times 10^{-5}$		
		(0.1 ~ 1) M $\Omega$	$1.8 \times 10^{-5}$		
		(1 ~ 10) M $\Omega$	$3.0 \times 10^{-5}$		
		(10 ~ 100) M $\Omega$	$2.5 \times 10^{-5}$		

## 404. Other DC &amp; LF Measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscopes	40421	(2 ~ 5) mV	$2.5 \times 10^{-2}$	Scopes calibrator /QECI-EL421
		(5 ~ 10) mV	$8.0 \times 10^{-3}$	
		(10 ~ 50) mV	$4.0 \times 10^{-3}$	
		(50 ~ 100) mV	$2.0 \times 10^{-3}$	
		(100 ~ 500) mV	$1.5 \times 10^{-3}$	
		(0.5 ~ 1) V	$2.0 \times 10^{-3}$	
		(1 ~ 5) V	$1.5 \times 10^{-3}$	
		(5 ~ 10) V	$2.0 \times 10^{-3}$	
		(10 ~ 50) V	$1.5 \times 10^{-3}$	
		(50 ~ 100) V	$2.0 \times 10^{-3}$	
		(2 ~ 5) ns	$5.0 \times 10^{-4}$	
		(5 ~ 50) ns	$1.0 \times 10^{-3}$	
		(50 ~ 500) ns	$1.0 \times 10^{-3}$	
		(0.5 ~ 5) $\mu$ s	$1.0 \times 10^{-3}$	
		(5 ~ 50) $\mu$ s	$1.0 \times 10^{-3}$	
Bandwidth	40421	(50 ~ 500) $\mu$ s	$1.0 \times 10^{-3}$	
		(0.5 ~ 5) ms	$1.0 \times 10^{-3}$	
		(5 ~ 50) ms	$1.0 \times 10^{-3}$	
		(50 ~ 500) ms	$1.0 \times 10^{-3}$	
		(0.5 ~ 5) s	$1.0 \times 10^{-3}$	
		600 mV		
		(0.05 ~ 100) MHz	$7.5 \times 10^{-2}$	
		(100 ~ 300) MHz	$1.2 \times 10^{-2}$	
		(300 ~ 600) MHz	$1.2 \times 10^{-2}$	
Volt/Current recorders	40424	(±)		Meter calibrator /QECI-EL424
		0 mV	$1.2 \mu\text{V}$	
		(0 ~ 10) mV	$6.5 \times 10^{-4}$	
		(10 ~ 100) mV	$8.5 \times 10^{-5}$	
		(0.1 ~ 1) V	$3.1 \times 10^{-5}$	
		(1 ~ 10) V	$2.0 \times 10^{-5}$	
		(10 ~ 100) V	$2.8 \times 10^{-5}$	
		(100 ~ 1 000) V	$2.8 \times 10^{-5}$	
		0 $\mu$ A	$24 \text{nA}$	
		(0 ~ 10) $\mu$ A	$1.2 \times 10^{-2}$	
		(10 ~ 100) $\mu$ A	$1.4 \times 10^{-3}$	
		(0.1 ~ 1) mA	$4.7 \times 10^{-4}$	
		(1 ~ 10) mA	$4.0 \times 10^{-4}$	
		(10 ~ 100) mA	$3.9 \times 10^{-4}$	
		(0.1 ~ 1) A	$4.3 \times 10^{-4}$	
		(1 ~ 20) A	$1.3 \times 10^{-3}$	

## 404. Other DC &amp; LF Measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Relay test sets	40425	60 Hz		Multimeter , Shunt
		(20 ~ 100) mV	$8.4 \times 10^{-4}$	/QECI-EL425
		(0.1 ~ 1) V	$2.8 \times 10^{-4}$	
		(1 ~ 10) V	$2.8 \times 10^{-4}$	
		(10 ~ 100) V	$2.8 \times 10^{-4}$	
		(100 ~ 1 000) V	$3.2 \times 10^{-4}$	
		60 Hz		
		(2 ~ 10) mA	$1.7 \times 10^{-3}$	
		(10 ~ 100) mA	$1.7 \times 10^{-3}$	
		(0.1 ~ 1) A	$2.5 \times 10^{-3}$	
		(1 ~ 10) A	$3.3 \times 10^{-3}$	
		(10 ~ 100) A	$1.3 \times 10^{-3}$	

## 501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature generators	50101	(-100 ~ 200) °C	0.9 °C	Data logger , Noble metal thermocouple
		(200 ~ 250) °C	1.5 °C	SPRT
		(200 ~ 1 100) °C	1.3 °C	/QECI-TE101
		(-196 ~ 200) °C	0.05 °C	/QECI-TE101-1
		(200 ~ 400) °C	0.1 °C	/QECI-TE101-2
		0 °C	0.01 °C	/QECI-TE101-3
		(-40 ~ 100) °C	0.06 °C	/QECI-TE101-4
		(100 ~ 400) °C	0.09 °C	
		(400 ~ 1 100) °C	0.8 °C	
Temperature indicators /recorders/controllers, temperature calibrators (Include sensors)	50102	(-196 ~ 400) °C	0.04 °C	SPRT, Noble metal thermocouple
		(400 ~ 800) °C	1.3 °C	Calibrator
		(800 ~ 1 100) °C	1.6 °C	/QECI-TE102
		(1 100 ~ 1 300) °C	2.6 °C	/QECI-TE102-1
		(-196 ~ 1 300) °C	0.47 °C	/QECI-TE102-2
		(-196 ~ 650) °C	0.15 °C	
Thermoelectric type (Exclude sensors)				
Resistance type (Exclude sensors)				

## 501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Glass thermometers; liquid-in-glass, Beckmann Glass thermometers	50103	(-40 ~ 360) °C	0.05 °C	SPRT /QECI-TE103
Resistance thermometers; IPRT, etc.	50104	(-196 ~ 400) °C	0.06 °C	SPRT /QECI-TE104
Thermal expansion thermometers; bimetal, gas or liquid types	50105	(-40 ~ 150) °C (150 ~ 400) °C	0.4 °C 0.7 °C	SPRT /QECI-TE105
Thermocouples; Base metal thermocouples	50106	(-196 ~ -40) °C (-40 ~ 200) °C (200 ~ 1 100) °C (1 100 ~ 1 300) °C (0 ~ 1 100) °C (1 100 ~ 1 300) °C	0.5 °C 0.3 °C 1.6 °C 2.7 °C 1.3 °C 2.4 °C	SPRT, Noble metal thermocouple /QECI-TE106-1 /QECI-TE106-2
Noble metal thermocouples				
Temperature transducers	50107	(-196 ~ 400) °C (400 ~ 1 100) °C (1 100 ~ 1 300) °C	0.3 °C 1.6 °C 2.8 °C	SPRT, Noble metal thermocouple /QECI-TE107

## 502. Non contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard radiation thermometers	50204	(0 ~ 200) °C (200 ~ 700) °C (700 ~ 1 000) °C	1.2 °C 1.6 °C 2.3 °C	Standard radiation thermometer /QECI-TE204

## 503. Humidity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Relative humidity hygrometers Polimer thin film hygrometers  Hair hygrometers	50302	(10 ~ 95) % R.H.	2.9 % R.H.	Dew point thermometer /QECI-HU302
		(-20 ~ 80) °C	0.6 °C	/QECI-HU302-1
		(20 ~ 90) % R.H.	4.4 % R.H.	
		(10 ~ 50) °C	0.7 °C	
Temperature humidity recorders; hygrothermograph, etc.	50304	(20 ~ 90) % R.H.	3.2 % R.H.	Dew point thermometer /QECI-HU304
		(10 ~ 80) °C	1.1 °C	
Transducers; dew-point/ relative humidity Ralative humidity transducers	50305			Dew point thermometer /QECI-HU305
		(10 ~ 50) % R.H.	2.6 % R.H.	
Humidity generators; two-pressure, two-temperature, flow mixing humidity gererator, constant temperature and humidity chamber, etc.	50306	(50 ~ 95) % R.H.	2.9 % R.H.	
		(10 ~ 30) % R.H.	2.2 % R.H.	Dew point thermometer /QECI-HU306
		(30 ~ 60) % R.H.	2.8 % R.H.	
		(60 ~ 80) % R.H.	4.0 % R.H.	
		(80 ~ 98) % R.H.	4.8 % R.H.	
		(-70 ~ 180) °C	0.9 °C	

## 701. Photometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Illuminance meters	70101	0.5 lx (0.5 ~ 1) lx (1 ~ 20 000)lx	$2.4 \times 10^{-2}$ $2.0 \times 10^{-2}$ $1.9 \times 10^{-2}$	Illuminance meters /QECI-PH701

## 901. Chemical analysis

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Breath alcohol analyzers	90101	0.030 % BAC (0.080 ~ 0.100) % BAC	$3.3 \times 10^{-2}$ $2.1 \times 10^{-2}$	Alcohol gas /QECI-CA101
Gas analyzers	90103			Standard gas
Oxygen ( $O_2$ )		(0 ~ 21.0) cmol/mol	$2.0 \times 10^{-2}$	/QECI-CA103
Carbon monoxide (CO)		(0 ~ 150.2) $\mu$ mol/mol	$2.0 \times 10^{-2}$	
Hydrogen sulfide ( $H_2S$ )		(0 ~ 13) $\mu$ mol/mol	$3.8 \times 10^{-2}$	
Methane ( $CH_4$ )		(13 ~ 30) $\mu$ mol/mol	$3.2 \times 10^{-2}$	
		(0 ~ 1.26) cmol/mol	$2.4 \times 10^{-2}$	
		(1.26 ~ 2.51) cmol/mol	$2.0 \times 10^{-2}$	
Carbon Dioxide ( $CO_2$ )		(0 ~ 2.54) cmol/mol	$2.0 \times 10^{-2}$	
Isobutylene (i- $C_4H_8$ )		100 $\mu$ mol/mol	1.0 $\mu$ mol/mol	
Isobutane (i- $C_4H_{10}$ )		(0 ~ 1) cmol/mol	$1.3 \times 10^{-2}$	
Hydrogen ( $H_2$ )		(0 ~ 2) cmol/mol	$1.1 \times 10^{-2}$	
Propane ( $C_3H_8$ )		1.06 cmol/mol	0.030 cmol/mol	
Nitric oxide (NO)		(0 ~ 10.2) $\mu$ mol/mol	$4.9 \times 10^{-2}$	
		(10.2 ~ 51.0) $\mu$ mol/mol	$2.9 \times 10^{-2}$	
Ammonia ( $NH_3$ )		50 $\mu$ mol/mol	2.4 $\mu$ mol/mol	

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & KS Q ISO/IEC 17025:2017

Q & Q Co., Ltd.

74, Seokgyesan-dan 1-gil, Sangbuk-myeon, Yangsan-si, Gyeongsangnam-do, Republic of Korea  
Phone : 82-55-912-0391, Fax : 82-55-912-0392, e-mail : qnqkoreays@naver.com

CALIBRATION

Valid To : Dec. 02, 2026.

Accreditation No : KC01-079

In recognition of the successful completion of the KOLAS evaluation process,  
accreditation is granted to this laboratory to perform the following calibrations

Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site
102. Linear dimension			10511	Measuring microscopes, profile projectors	Y			
10206	Dial/cylinder gauge testers	N	10514	Taper plug gauges	N			
10207	Doctor blades	N	106. Various dimensional					
10209	End bars	N	10601	Inside/outside/gear tooth calipers, caliper gauges	Y			
10212	Film applicators	N	10603	Cylinder/bore gauges	Y			
10213	Gap gauges	N	10604	Depth gauges, depth micrometers	Y			
10214	Gauge blocks, by comparison	N	10605	Dial/digital gauges	Y			
10216	Height gauges/measuring machines	Y	10609	Micro indicators, test indicators	Y			
10220	Standard measuring machines	Y	10611	3-points, micrometers	Y			
10223	Electronic micrometers	N	10612	Inside micrometers	Y			
10224	Height micrometers, riser blocks	N	10613	Outside micrometers	Y			
10229	Radius gauges	N	10617	Standard sieves	N			
10233	Taper thickness gauges	N	10620	Welding gauges	Y			
10234	Ultrasonic thickness gauges	Y						
10236	Coating thickness testers	Y						
103. Angle								
10304	Bevel protractors	N						
10311	Plate/square/electric levels	N						
10318	Squareness testers, right angle testers	N						
10320	Precision squares	N						
104. Form								
10404	Optical flats	N						
10405	Optical parallels	N						
10406	Parallel blocks	N						
10407	Precision surface plates	Y						
10415	Test bars	N						
105. Complex geometry								
10502	Bench centers	N						
10503	Contact coordinate measuring machines	Y						
10504	Non-contact coordinate measuring machines	Y						

Note

- This laboratory provides calibration services in permanent standard laboratory and at on-site.
- Laboratory conducts on-site calibration should meet requirements of KOLAS-SR-007.
- On-site calibration is allowed to items with marking 'Y', not allowed to items with marking 'N'.
- Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k=2$ . It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.
- Due to the calibration environment such as reference standards or customers' facilities, it is note that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.

## 102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dial/cylinder gauge testers Q & Q Co., Ltd.	10206	(0 ~ 25) mm	$\sqrt{(0.20 \text{ } \mu\text{m})^2 + (2.9 \times 10^{-6} \times l_0)^2}$	Gauge block, Electronic micrometer /QYCI-LE206
Doctor blades	10207	(0 ~ 10) mm	1.1 $\mu\text{m}$	Height micrometer, Electronic micrometer /QYCI-LE207
End bars	10209	(25 ~ 500) mm (500 ~ 1 000) mm	$\sqrt{(0.33 \text{ } \mu\text{m})^2 + (3.0 \times 10^{-6} \times l_0)^2}$ $\sqrt{(1.3 \text{ } \mu\text{m})^2 + (2.9 \times 10^{-6} \times l_0)^2}$	Gauge block, Electronic micrometer /QYCI-LE209
Film applicators	10212	(0 ~ 10) mm	1.1 $\mu\text{m}$	Height micrometer, Electronic micrometer /QYCI-LE212
Gap gauges	10213	(2 ~ 200) mm	$\sqrt{(1.4 \text{ } \mu\text{m})^2 + (2.7 \times 10^{-6} \times l_0)^2}$	Height micrometer, Electronic micrometer /QYCI-LE213
Gauge blocks, by comparison	10214	(0.5 ~ 100) mm	$\sqrt{(80 \text{ nm})^2 + (1.2 \times 10^{-6} \times l_0)^2}$	Gauge block, Gauge block comparator /QYCI-LE214
Height gauges/measuring machines	10216			Gauge block /QYCI-LE216
Height gauges		(0 ~ 1 500) mm	$\sqrt{(10 \text{ } \mu\text{m})^2 + (3.0 \times 10^{-6} \times l_0)^2}$	
Height measuring machines		(0 ~ 600) mm	$\sqrt{(0.5 \text{ } \mu\text{m})^2 + (0.3 \times 10^{-5} \times l_0)^2}$	
Standard measuring machines	10220	(0 ~ 500) mm	$\sqrt{(0.20 \text{ } \mu\text{m})^2 + (2.8 \times 10^{-6} \times l_0)^2}$	Gauge block /QYCI-LE220
Electronic micrometers	10223	$\pm 50 \text{ } \mu\text{m}$ $\pm 50 \text{ } \mu\text{m} \sim \pm 2 \text{ mm}$	0.1 $\mu\text{m}$ 1 $\mu\text{m}$	Gauge block /QYCI-LE223
Height micrometers	10224			Gauge block, Electronic micrometer /QYCI-LE224
Block calibration		(0 ~ 310) mm	$\sqrt{(0.74 \text{ } \mu\text{m})^2 + (3.0 \times 10^{-6} \times l_0)^2}$	/QYCI-LE224
Head calibration		(0 ~ 20) mm	0.8 $\mu\text{m}$	/QYCI-LE224-1
Riser blocks		150 mm 300 mm 600 mm	0.9 $\mu\text{m}$ 1.1 $\mu\text{m}$ 2.2 $\mu\text{m}$	
Parallelism			0.8 $\mu\text{m}$	
Radius gauges	10229	(0.03 ~ 100) mm	2 $\mu\text{m}$	Non-contact coordinate measuring machine /QYCI-LE229

## 102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Taper thickness gauges	10233	(0 ~ 90) mm	2 $\mu\text{m}$	Non-contact coordinate measuring machine /QYCI-LE233
Ultrasonic thickness gauges	10234	(0 ~ 200) mm	$\sqrt{(10 \mu\text{m})^2 + (3.9 \times 10^{-6} \times l_0)^2}$	Ultrasonic specimen /QYCI-LE234
Coating thickness testers	10236	(0 ~ 3.9) mm	$\sqrt{(1.8 \mu\text{m})^2 + (2.7 \times 10^{-6} \times l_0)^2}$	Thickness specimen /QYCI-LE236

## 103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Bevel protractors	10304			Non-contact coordinate measuring machine
Angle accuracy		(0 ~ 180) $^\circ$	1'	/QYCI-AN304
Straightness		(0 ~ 300) mm	1.3 $\mu\text{m}$	
Plate/square/electric levels	10311			Level comparator,
Angle(Bubble Tube Type)		$\pm 20.6''$	$\sqrt{(0.8'')^2 + (0.3 \times 10^{-3} \times r_0)^2}$	Electronic micrometer /QYCI-AN311
Flatness of Base		(0 ~ 300) mm	1.0 $\mu\text{m}$	
Squareness		(0 ~ 300) mm	8.3 $\mu\text{m}/\text{m}$	
Squareness testers	10318			Standard cylindrical square, Electronic micrometer
		(0 ~ 490) mm	$\sqrt{(1.4 \mu\text{m})^2 + (3.0 \times 10^{-6} \times l_0)^2}$	/QYCI-AN318
Precision squares	10320			Standard cylindrical square,
Squareness		(0 ~ 490) mm	$\sqrt{(1.3 \mu\text{m})^2 + (3.0 \times 10^{-6} \times l_0)^2}$	Electronic micrometer
Parallelism		(0 ~ 500) mm	1.7 $\mu\text{m}$	/QYCI-AN320
Straightness		(0 ~ 500) mm	1.9 $\mu\text{m}$	

## 104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical flats	10404	(Ø 10 ~ Ø 100) mm	$\sqrt{(23 \text{ nm})^2 + (4.3 \times 10^{-7} \times d_0)^2}$	Optical flat /QYCI-LE404
Optical parallels	10405	(Ø 10 ~ Ø 50) mm	0.04 μm 0.08 μm	Optical flat, Gauge block comparator /QYCI-LE405
Parallel blocks	10406	(0 ~ 500) mm	1.9 μm 1.9 μm 1.9 μm	Electronic micrometer /QYCI-LE406
Precision surface plates	10407	(0.09 ~ 1.44) m <sup>2</sup> (1.44 ~ 4.84) m <sup>2</sup> (4.84 ~ 9) m <sup>2</sup> (9 ~ 17.5) m <sup>2</sup>	1.3 μm 1.9 μm 2.6 μm 3.1 μm	Electronic Level /QYCI-LE407
Test bars	10415	(0 ~ 500) mm (Ø 10 ~ Ø 100) mm	$\sqrt{(1.3 \text{ μm})^2 + (4.1 \times 10^{-6} \times l_0)^2}$ 1.0 μm 1.0 μm 3"	Gauge block, Electronic micrometer, Outside micrometers /QYCI-LE415

## 105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Bench centers Height difference of both the center Flatness of the bed side	10502	(0 ~ 500) mm	2.6 $\mu\text{m}$  2.3 $\mu\text{m}$	Test bar, Electronic micrometer /QYCI-LE502
Contact coordinate measuring machines Indicating accuracy Squareness Straightness	10503	(0 ~ 1 500) mm  (0 ~ 1 000) mm  (0 ~ 500) mm	$\sqrt{(2.7 \mu\text{m})^2 + (5.5 \times 10^{-6} \times l_0)^2}$  4.0 $\mu\text{m}$  1.7 $\mu\text{m}$	Gauge block, Precision square, Straight edge, Step Gauge /QYCI-LE503
Non-contact coordinate measuring machines Indicating accuracy Squareness	10504	(0 ~ 600) mm  (0 ~ 450) mm	$\sqrt{(0.40 \mu\text{m})^2 + (2.8 \times 10^{-6} \times l_0)^2}$  2"	Standard scale, Angle gauge block /QYCI-LE504
Measuring microscopes, Profile projectors Feed accuracy of workstage Squareness Magnification Error Angle division accuracy	10511	(0 ~ 300) mm  2.6 $\mu\text{m}$  2.4 $\times 10^{-4}$  1.1'	$\sqrt{(0.51 \mu\text{m})^2 + (2.3 \times 10^{-6} \times l_0)^2}$  $\sqrt{(0.51 \mu\text{m})^2 + (2.3 \times 10^{-6} \times l_0)^2}$  $\sqrt{(0.51 \mu\text{m})^2 + (2.3 \times 10^{-6} \times l_0)^2}$	Standard scale, Square /QYCI-LE511 /QYCI-LE511-1
Taper plug gauges Taper half angle Small-end diameter Step diameter Big-end diameter Gauge length Step length	10514	(0 ~ 65)°  (Ø 2 ~ Ø 100) mm  (Ø 2 ~ Ø 100) mm  (Ø 2 ~ Ø 100) mm  (0 ~ 250) mm  (0 ~ 150) mm	2"  $\sqrt{(1.2 \mu\text{m})^2 + (3.0 \times 10^{-6} \times m_0)^2}$  $\sqrt{(1.5 \mu\text{m})^2 + (3.0 \times 10^{-6} \times d_1)^2}$  $\sqrt{(2.1 \mu\text{m})^2 + (3.0 \times 10^{-6} \times d_2)^2}$  $\sqrt{(0.80 \mu\text{m})^2 + (2.9 \times 10^{-6} \times l_0)^2}$  $\sqrt{(0.80 \mu\text{m})^2 + (2.9 \times 10^{-6} \times l_0)^2}$	Gauge block, Electronic micrometer, Outside micrometers, Height micrometer /QYCI-LE514

## 106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inside/outside/gear tooth calipers, caliper gauges Inside/Outside calipers Gear tooth calipers Tooth height scale Tooth thickness scale Combine accuracy Caliper gauges	10601	(0 ~ 1 500) mm	$\sqrt{(10 \text{ } \mu\text{m})^2 + (3.0 \times 10^{-6} \times l_0)^2}$	Gauge block, Step gauge, Surface plate /QYCI-LE601 /QYCI-LE601-1 /QYCI-LE601-2
Cylinder/Bore gauges	10603	(0 ~ 5) mm	0.3 $\mu\text{m}$	Dial gauge tester /QYCI-LE603
Depth gauges, Depth micrometers Depth gauges Depth micrometers Dial depth gauges	10604	(0 ~ 500) mm	$\sqrt{(10 \text{ } \mu\text{m})^2 + (2.8 \times 10^{-6} \times l_0)^2}$	Gauge block, Surface plate /QYCI-LE604 /QYCI-LE604-1 /QYCI-LE604-2
Dial/digital gauges	10605	(0 ~ 1) mm	$\sqrt{(0.30 \text{ } \mu\text{m})^2 + (3.9 \times 10^{-6} \times l_0)^2}$	Dial gauge tester, Gauge block /QYCI-LE605
		(1 ~ 25) mm	$\sqrt{(1.0 \text{ } \mu\text{m})^2 + (3.9 \times 10^{-6} \times l_0)^2}$	
Micro indicators, Test indicators Micro indicators Test indicators	10609	$\pm 1 \text{ mm}$ (0 ~ 2) mm	0.3 $\mu\text{m}$ 0.3 $\mu\text{m}$	Dial gauge tester /QYCI-LE609 /QYCI-LE609-1
3-points, Micrometers	10611	( $\emptyset 3$ ~ $\emptyset 6$ ) mm ( $\emptyset 6$ ~ $\emptyset 100$ ) mm	1.1 $\mu\text{m}$ 2 $\mu\text{m}$	Cylindrical ring gauge /QYCI-LE611
Inside micrometers Inside micrometers Tubular inside micrometers	10612	(5 ~ 300) mm (50 ~ 600) mm (600 ~ 1 500) mm	$\sqrt{(1.0 \text{ } \mu\text{m})^2 + (3.0 \times 10^{-6} \times l_0)^2}$	Gauge block, Gauge block accessory /QYCI-LE612 /QYCI-LE612-1

## 106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Outside micrometers	10613	(0 ~ 100) mm	$\sqrt{(1.0 \text{ } \mu\text{m})^2 + (2.9 \times 10^{-6} \times l_0)^2}$	Gauge block, Optical flat, Cylindrical plug gauge /QYCI-LE613 /QYCI-LE613-1
		(100 ~ 500) mm	$\sqrt{(1.0 \text{ } \mu\text{m})^2 + (3.0 \times 10^{-6} \times l_0)^2}$	
		(500 ~ 1 000) mm	$\sqrt{(2.0 \text{ } \mu\text{m})^2 + (3.0 \times 10^{-6} \times l_0)^2}$	
		(1 ~ 85) mm	$\sqrt{(1.0 \text{ } \mu\text{m})^2 + (4.6 \times 10^{-6} \times l_0)^2}$	
V-Anvil micrometers				
Standard sieves	10617			Non-contact coordinate measuring machine /QYCI-LE617
		Wire diameter	3 $\mu\text{m}$	
		Sieve size	4 $\mu\text{m}$	
		Standard plate sieve		
Length of the hole center		Hole diameter	3 $\mu\text{m}$	
		(0 ~ 130) mm	3 $\mu\text{m}$	
Welding gauges	10620	(0 ~ 160) mm	3 $\mu\text{m}$	Non-contact coordinate measuring machine, Gauge block /QYCI-LE620
		Height, Depth	0.3 mm	
		Scale	0.3 mm	
		Fillet Welding Height	0.3 mm	
		Taper thickness	0.2 mm	
		Angle	0.3°	